King's Research Portal

DOI:

[10.1016/j.juro.2016.01.131](https://doi.org/10.1016/j.juro.2016.01.131)

Document Version

Peer reviewed version

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

Aydin, A., Shafi, A. M. A., Shamim Khan, M., Dasgupta, P., & Ahmed, K. (2016). Current Status of Simulation and Training Models in Urological Surgery: A Systematic Review. *The Journal of urology*, 196(2), 312-320. <https://doi.org/10.1016/j.juro.2016.01.131>

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Author's Accepted Manuscript

Current Status of Simulation and Training Models in Urological Surgery: A Systematic Review

Abdullatif Aydin , Ahmed M.A. Shafi , Muhammad Shamim Khan , Prokar Dasgupta , Kamran Ahmed

PII: S0022-5347(16)03683-1
DOI: [10.1016/j.juro.2016.01.131](https://doi.org/10.1016/j.juro.2016.01.131)
Reference: JURO 13594

To appear in: *The Journal of Urology*
Accepted Date: 10 January 2016

Please cite this article as: Aydin A, Shafi AMA, Khan MS, Dasgupta P, Ahmed K, Current Status of Simulation and Training Models in Urological Surgery: A Systematic Review, *The Journal of Urology*® (2016), doi: 10.1016/j.juro.2016.01.131.

DISCLAIMER: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our subscribers we are providing this early version of the article. The paper will be copy edited and typeset, and proof will be reviewed before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to The Journal pertain.

Embargo Policy

All article content is under embargo until uncorrected proof of the article becomes available online.

We will provide journalists and editors with full-text copies of the articles in question prior to the embargo date so that stories can be adequately researched and written. The standard embargo time is 12:01 AM ET on that date. Questions regarding embargo should be directed to jumedia@elsevier.com.

**Current Status of Simulation and Training Models in Urological Surgery: A
Systematic Review**

Authors:

Abdullatif Aydin BSc (Hons), MBBS

Ahmed M.A. Shafi BSc (Hons)

Muhammad Shamim Khan FRCS (Urol), FEBU, OBE

Prokar Dasgupta MSc, MD, FRCS (Urol), FEBU

Kamran Ahmed MRCS, PhD

Affiliations:

MRC Centre for Transplantation, King's College London

Department of Urology, Guy's and St. Thomas' NHS Foundation Trust

King's Health Partners, London, United Kingdom

Correspondence:

Kamran Ahmed MBBS, MRCS, PhD

NIHR / TUF Academic Clinical Lecturer

Specialist Registrar in Urology

King's College London, Guy's Hospital, London

Ph: +44 (0)7939319670

Email: kamran.ahmed@kcl.ac.uk

Key words: assessment, endourology, laparoscopy, models, robotic, simulation, systematic review, training, urology

Word count of text: 3982

Word count of abstract: 243

ABSTRACT

Purpose: Increased awareness of patient safety, advances in surgical technology and reduced working times have led to the adoption of simulation-enhanced training. However, simulators available need to be scientifically evaluated before integration into curricula. The aim of this study is to identify the currently available training models for urological surgery, their status of validation and evidence behind each model.

Materials and Methods: Medline, EMBASE and Cochrane library databases were searched for English language articles published between 1990 and 2015, describing urological simulators and/or validation studies of these models. All studies were assessed for level of evidence and each model was subsequently awarded a level of recommendation, using a modified Oxford Centre for Evidence-Based Medicine classification, adapted for education by the European Association of Endoscopic Surgeons.

Results: A total of 91 validation studies were identified pertaining to training models in endourology (n=63), laparoscopic surgery (n=17), robot-assisted surgery (n=8) and open urological surgery (n=6), with a total of 55 models. Of the included studies, six were classified Level 1b, nine Level 2a, 39 Level 2b and 19 Level 2c. Amongst all the training models, the URO-Mentor was the only model to receive a Level of Recommendation of 1.

Conclusions: UroSimulation is a growing field and increasing numbers of models are being produced. However, there are still too few validation studies with high level of evidence demonstrating the transferability of skills. Nevertheless, efforts should be made to utilise the currently available models in curriculum-based training programmes.

INTRODUCTION

Surgery as a craft has traditionally been learnt via an apprenticeship scheme, in which the famous slogan, “see one, do one, teach one” was coined, describing how surgical skills were attained for many decades¹. This method of training produced highly skilled surgeons for a number of generations. However, with increased awareness of patient safety, reduced working hours and financial constraints in healthcare organisations, this model has presented challenges for trainees to obtain the required level of competency².

Further challenges have risen with the development of minimally invasive techniques, largely associated with steeper learning curves. With growing realisation that a large part of the procedural learning curve does not require patients for skill acquisition, but can be learnt on training models, there has been a boom in the production of training models³. This increase has brought about the need to scientifically evaluate these models to establish their educational value and role in training. Thus, increasing number of validation studies are being conducted to investigate the usefulness of simulators.

The aim of this study is to identify the currently available training models for urological surgery and their status of validation. It also aims to evaluate the level of evidence (LoE) behind each training model and, thereby, establish a level of recommendation (LoR).

METHODS

This study was performed using the guidelines set out by the Preferred Reporting Items for Systematic reviews And meta-analysis (PRISMA) statement (Figure 1) ⁴.

Study Eligibility Criteria

Original research articles describing validation and use of urology training models and simulators were included in the review. Studies addressing basic surgical skills were excluded, as were those that only described the use of models without a validation process. Abstracts with insufficient information and non-English articles were also excluded.

Information Sources and Search

A broad search was performed on Medline, EMBASE and Cochrane library databases, between January 1990 and December 2015. Search terms included “simulation in urology” and “simulation training in urology”, which allowed for the majority of articles to be found. A further procedure-specific search was performed using “TURP”, “TURBT”, “Nephrectomy”, “HoLEP”, “PVP”, “PCNL”, “Laser”, “diode”, “GreenLight” and “robotic” followed by “simulation” or “training”, to supplement the identified studies.

Study Selection and Data Collection

After meeting the inclusion criteria, articles were retrieved in full and titles and abstracts were examined. Abstracts from conferences were also included, if sufficient information could be extracted. Duplicates were removed. Full-text review further excluded studies, which were not validation or educational impact studies. Potentially relevant articles found

in the references of included articles were also retrieved and made subject to the inclusion/exclusion criteria.

Data Items

Selected data was extracted from each study including the name of model, institution/manufacture, type of validation, number of participants and demographics. Models and simulators were classified into the following categories: bench, augmented reality (AR), virtual reality (VR), animal and cadaveric models. Results were tabulated and studies concerning each simulator were grouped together. The type of validities were classified according to the definitions of McDougall⁵ and van Nortwick et al.⁶ (Figure 2). LoE for each study and LoR for each model was awarded using a modified educational Oxford Centre for Evidence-Based Medicine (OCEBM) LoE and LoR classification system, as adapted by the European Association of Endoscopic Surgery⁷ (Supplementary Tables).

RESULTS

In total, 5163 potentially relevant studies were identified. Upon review and examination of the full texts, 91 of the initially retrieved studies met the inclusion criteria (Figure 1). Results were categorised into endourology, laparoscopic urology, robot-assisted urological surgery and open urological surgery (Tables 1-5). In cases where studies failed to demonstrate the aimed validity, these were indicated with a strike through the text and where studies did not provide adequate information, these were hyphenated in the relevant sections of the tables. In instances where LoR could not be awarded, these were marked with not-applicable (N/A).

Endourology

Sixty-three studies were identified concerning training models for different endourological procedures, most of which were for ureterorenoscopy followed by urethrocystoscopy and transurethral resection, respectively.

Urethrocystoscopy

Eighteen articles were identified describing validated models utilised in flexible and rigid cystoscopy training (Table 1), with eleven using the URO Mentor VR simulator (Simbionix, Israel) alone. The URO Mentor has demonstrated face (n=4), content (n=2), construct A (n=6) and construct B validities (n=6) for urethrocystoscopy, and was awarded LoR of 1. This VR simulator was also evaluated when used in conjunction with a bench model. Gettman et al.⁸ developed a standardised cystoscopy curriculum using the URO Mentor and showed significant improvements with additional simulator time amongst 10 novices. A

complementary study ⁹ with 80 subjects demonstrated construct validity of the model; while a randomised controlled trial (RCT) ¹⁰ of 100 subjects demonstrated transfer of skills from the VR simulator to the operating room.

Fresh frozen cadavers (FFC) were utilised for rigid and flexible cystoscopy by two studies ^{11 12} and received a LoR of 2. Bowling et al. ¹¹ demonstrated construct B validity amongst 29 obstetric residents whilst Ahmed et al. ¹² described the British Association of Urological Surgeons (BAUS) Human Cadaver Training Programme and demonstrated face and content validity amongst 75 participants and 27 experts for a number of procedures including rigid and flexible cystoscopy, bladder biopsy and intravesical injection of botulinum toxin.

Ureterorenoscopy

Twenty-one studies were identified describing validated training models for ureterorenoscopy (Table 1), 11 of which described the URO Mentor once again. This VR simulator has demonstrated face (n=4), content (n=2), construct A (n=5), construct B (n=4), concurrent (n=1) and predictive validities (n=2) for ureterorenoscopy, and was awarded a LoR of 2.

Another more commonly used model is the Uro-Scopic Trainer (Limbs and Things, Bristol, UK). This bench model was evaluated by four studies and demonstrated face (n=1), construct A (n=3), construct B (n=1) and concurrent validities (n=1). The model received a LoR of 2. Two studies evaluated the combined use of these models ^{13 14} and demonstrated face and construct B validation, amongst 16 residents and 32 medical students, respectively.

Matsumoto et al.¹⁵ also demonstrated Construct A validity of a low-fidelity model (cost: \$20 CAN), consisting of a penrose drain, inverted cup, moulded latex and two straws, amongst 40 medical students. It, too, received a LoR of 2. This RCT compared the low fidelity model to the Uro-Scopic Trainer, a higher fidelity bench model (cost: \$3700 CAN), and didactic lectures. Statistically significant improvements were seen in the two groups that had simulation training compared to a group that only received didactic lectures. However no differences were observed between the low and high fidelity group.

Several other bench models were also validated, including the Scope Trainer (Mediskills, Northampton, UK), Adult Ureteroscopy Trainer (Ideal Anatomic Modelling, Holt, Michigan, USA), Endo-Urologie Modell (Karl Storz GmbH, Tuttlingen, Germany) and Cook URS Model (Cook Medical, Bloomington, IN, USA), all of which received a LoR of 3.

Soria et al.¹⁶ developed a three-stage curriculum whereby 40 participants were first taught theoretical knowledge and then utilised the ETXY-Uro Adam (ProDelphus, Olinda, Brazil) to perform urethrocystoscopy, ureteral orifice cannulation and a semi-rigid ureteroscopy case. This was followed by laser lithotripsy and basket removal of stones on a porcine renoureteral unit. Finally, participants repeated task 1 on live porcine. The authors demonstrated face, content, construct A and construct B validity of their curriculum, receiving a LoE of 2b and LoR of 3.

Face and content validity of FFCs were demonstrated¹², but did not have sufficient LoE to receive a LoR.

Percutaneous renal surgery

Eight articles were identified for validated models utilised in the training of percutaneous renal surgery, describing a total of four training models (Table 1). The PERC Mentor (Simbionix, Israel) was evaluated by four studies and demonstrated face (n=2), content (n=2), construct A (n=2), construct B (n=2) and predictive validities (n=1). The simulator received a LoR of 2. Another study¹⁷ used the PERC Mentor in conjunction with live porcine and demonstrated the content validity of both modalities.

The only identified bench model was the C-arm Trainer (SimPORTAL, University of Minnesota, MN, USA). The developers assessed and demonstrated the face and content validity of this new model. Two studies reported the use of porcine kidney for PCNL while. Zhang et al.¹⁸ reported the face validity of using porcine kidneys amongst 42 urologists whilst Hammond et al.¹⁹ placed the porcine kidney into a chicken carcass to represent the posterior tissue layers in humans and demonstrated face validity of this model amongst urology residents. All three models received a LoE of 4 and failed to receive a LoR.

Transurethral resection of the prostate (TURP)

Sixteen validation studies were identified for TURP training models (Table 2). Three training models equally received a LoR of 2. The Bristol TURP Trainer (Limbs and Things), SurgicalSIM TURP Trainer (METI, Seattle, WA, USA) and the UroSim TURP module (VirtaMed, Zurich, Switzerland) all demonstrated face, content and construct B validities. The PelvicVision VR Simulator (Melereit Medical AB, Linköping, Sweden) also demonstrated face, content and construct validity, but received a LoR of 3. Of these, the Bristol TURP Trainer is the only bench model, as opposed to the remaining VR simulators.

The TURP Simulator described and content validated by Ballaro et al.²⁰ was developed over a decade ago but has not been reported in the literature since or commercially available. The Uro Trainer (Karl Storz GmbH, Tutlingen, Germany) is also no longer commercially available and the name is, instead, used synonymously with custom-made UroSim, produced for Karl Storz.

Ahmed et al.¹² also described the TURP simulation with FFCs as part of the BAUS Human Cadaver Training Programme and demonstrated face and content validity.

Transurethral resection of bladder tumours (TURBT)

Three training models were identified for TURBT training (Table 2). The Uro Trainer VR simulator (Karl Storz GmbH) failed to demonstrate face and content validity in the study conducted by Schoutt et al.²¹ whilst other studies demonstrated content (n=1), construct A (n=3) and construct B (n=1) validities of the simulator. However, as mentioned, this is no longer available and the TURBT module on the UroSim remains to be validated. Khan et al.²² utilised the Bristol TURBT Trainer (Limbs and Things) and also demonstrated face, content and construct B validity. Similarly, the Simbla TURBT Simulator was also evaluated by de Vries et al.²³ and demonstrated face, content and construct B validity amongst 76 participants consisting of novices, intermediaries and experts. All three models received a LoR of 3.

Laser procedures of the prostate

A total of six studies were identified for models utilised in training for various laser prostate therapies (Table 2). These consisted of three VR simulators and one bench-top model. Two

simulation models are available for Holmium laser enucleation of the prostate (HoLEP); the UroSim (VirtaMed) and HoLEP Simulator (Kansai Medical University, Japan). The former has demonstrated face, content and construct validity whilst the latter has only demonstrated face and content.

The GreenLight Simulator (American Medical Systems Inc., Minnetonka, MN, USA) is the only available simulator for GreenLight photoselective vaporisation of the prostate (PVP). Three studies evaluated this simulator and demonstrated face (n=2), content (n=1), construct A (n=2) and construct B validities (n=3). The MyoSim (VirtaMed) also demonstrated construct B validity by Angulo et al.²⁴ for diode laser PVP. The GreenLight Simulator was the only trainer to receive a LoR of 2, with the remaining scoring LoR of either ≤ 3 or N/A.

Laparoscopic Urology

Sixteen studies were identified describing validated models for three laparoscopic procedures (Table 3). The most commonly used modality was animal models.

Partial/Radical Nephrectomy

A total of eight studies described four different models for laparoscopic partial and/or radical nephrectomy. The ProCedicus MIST Nephrectomy VR Simulator (Mentice, Gothenburg, Sweden) was the most thoroughly evaluated model, with two studies demonstrating its face, content and construct B validities. However, Wijn et al.²⁵ reported failure of the trainer to demonstrate construct validity in a cohort relatively higher in number. Nevertheless, this VR simulator received an overall LoR of 2. Similarly, the Partial

Nephrectomy dry-lab model (University of California, USA) also received a LoR of 2, with three studies demonstrating face (n=1), content (n=1) and construct B (n=2) validation.

Pyeloplasty

Four studies were identified for Laparoscopic Pyeloplasty, describing different bench (n=2) and animal (n=2) models, all of which received LoR of ≤ 3 . Of note, Poniatowsky et al.²⁶ developed and validated a disposable, low-cost, high-fidelity, physical renal pelvis/ureter tissue analogue model for pyeloplasty training. The authors demonstrated face, content and construct validity amongst 31 participants. Jiang et al.²⁷ used a chicken crop and oesophagus model to simulate the human renal pelvis and ureter, respectively. This study demonstrated construct B validity of this model amongst 15 participants, where experienced participants outperformed intermediate and novices.

Ureteral Re-implantation

Tunitsky et al.²⁸ developed a dry-lab model for minimally invasive ureteral re-implantation from hydrogel. The authors demonstrated face, content and construct B validity amongst 20 subjects. The model received a LoR of ≤ 3 .

Urethrovesical Anastomosis

Three studies were identified for urethro-vesical anastomosis models (UVA), all of which received LoR of ≤ 3 . The RCT by Sabbagh et al.²⁹ evaluated a latex task-specific simulator for UVA and demonstrated its face and predictive validities. Anaesthetised pigs were used to assess performance, where the groups who trained on the model outperformed the control arm. The authors also demonstrated the face validity of the live porcine model. A number of

other studies have also assessed the used of various animal models, including pig intestine and chicken models, and demonstrated construct B validity.

Robot-Assisted Urological Surgery

Eight studies were identified describing procedure-specific and validated training models for robot-assisted urological surgery (Table 4). The most commonly used robotic simulator was the da Vinci Skills Simulator (dVSS; Intuitive Surgical, Sunnyvale, CA, USA).

Robot-assisted partial nephrectomy (RAPN)

Three validation studies were identified for two RAPN models. Hung et al.³⁰ developed a partial nephrectomy model from porcine kidney and a Styrofoam ball, and established face, content and construct B validity amongst 46 participants, using the dVSS. Concurrent and predictive validities of the model were also established in a second study, amongst 24 participants. This model was awarded a LoR of 2. The same authors also developed a procedure-specific AR training platform on the dV-Trainer (Mimic Technologies, Seattle, WA, USA) for RAPN. The AR component includes surgical footage embedded with interactive exercises to teach relevant surgical anatomy, steps of the operation and advanced surgical skills such as tissue retraction, cutting and energy use. This is combined with a renorrhaphy VR exercise. The authors established face, content, construct B and concurrent validities. This platform received a LoR of 3.

Robot-assisted radical prostatectomy (RARP)

Three validation studies described two training models for UVA, an essential step in RARP, and received a LoR of 3. The Hands-on Surgical Training (HoST) is an AR system on the

Robotic Surgical Simulator (RoSS; Simulated Surgical Systems; Williamsville, NY, USA). Chowriappa et al.³¹ performed a prospective RCT and demonstrated face and concurrent validities, amongst 52 participants. Although the study is of high quality (Level 1b), the lack of further studies on the model limit its LoR. The Tube 3 module on the dV-Trainer also simulates UVA and has demonstrated face, content and construct B validity amongst 20 participants³². Kim et al.³³ recruited 11 novices and trained them using this module over seven sessions. The participants then performed UVA on synthetic materials and, later, on a UVA model, thereby demonstrating a level of concurrent and predictive validities.

Two studies were identified, describing models used in full procedural simulation for RARP. A chicken genitourinary model was developed and used to perform a full RARP and demonstrated face, content and construct B validity amongst 20 participants³⁴. The European Association of Urology (EAU) Robotic Urologic Section (ERUS) has designed and developed a structured training program and curriculum in urology focusing on RARP. The simulation component of the curriculum includes training on VR, dry-lab, animal and cadaveric models. Volpe et al.³⁵ established the face validity of the ERUS Curriculum amongst ten robotic fellows. Currently, the curriculum has a LoR of 4.

Open Urological Surgery

Six studies were identified describing training models for 15 open urological procedures (Table 5). The majority of the procedures were performed on FFCs, conducted as part of the BAUS training programme¹². The authors demonstrated face and content validity of FFCs for common and emergency urological operations, amongst 75 residents and 27 experts. This study was awarded a LoE of 4. Similarly, Cabello et al.³⁶ demonstrated the use of thiel-

embalmed cadavers in Renal transplantation and demonstrated face validity amongst 28 subjects. Several bench models were also described and validated for suprapubic catheterisation (SPC; n=3) and vasovasostomy (n=1). Of these, the Uro-Emerge SPC model and the silicone tube and rat vas deferens vasovasostomy models both demonstrated construct A and predictive validities, achieving a LoE of 2c and LoR of 4. Due to the nature of the studies and lack of supportive data, it was not possible to make recommendations for the remaining models in this section.

Non-technical Skills

Two individual concepts were identified for integrated non-technical skills training in urology, both of which received a LoR of 2. First is full immersion simulation (Imperial College, London, UK), a low-fidelity 360° inflatable operating environment. Brunckhorst et al.¹⁴ conducted a RCT where the randomised arm received ureteroscopy and non-technical skills training through a validated curriculum and assessed both arms. The randomised group outperformed the control arm in all aspects including non-technical skills, thereby demonstrating construct A validity. Furthermore, the authors confirmed a strong correlation between technical and non-technical skills³⁷. Brewin et al.³⁸ also used full immersion simulation for TURP and non-technical skills and demonstrated face, content and construct B validity, with a Kirkpatrick Level 1 Evidence for educational impact. Full immersion simulation received.

Two studies^{39 40} utilised high fidelity simulation within a simulated laparoscopic Operating Room (OR; Storz, Tuttlingen, Germany) using a high-fidelity SimMan3G mannequin (Laerdal, Wappingers Falls, NY) and Partial nephrectomy renal model (Table 3). The authors executed

a partial nephrectomy scenario with complications. High-fidelity OR team training demonstrated face (n=2), content (n=2) and construct B validation in both technical skills (n=2) and non-technical skills (n=1).

DISCUSSION

Surgical simulation has seen an exponential growth in the last couple of decades, with increasing number of new models being developed and validated. This study has demonstrated that simulation-based urological training has made considerable progress. High numbers of procedure-specific models have been developed for endourology and also a select few for laparoscopic and robot-assisted urology, which previously concentrated on generic skills acquisition⁴¹. Four Level 1b studies were identified in endourology^{10 11 15 42} and two in robot-assisted urology^{31 43}. In contrast, open urological surgery has had a limited number of simulator production and subsequent validation, which may owe to the nature of the surgery, making it difficult to replicate.

Based on the currently available data, simulation-based urological training should be adopted by healthcare organisations, through utilisation of validated models. Efforts should be made to identify the best aspects of each training model and procedure-specific simulation curricula should be developed and validated, employing different modalities. Several generic skills curricula have been reported in the literature including the validated Fundamentals of Laparoscopic Surgery (FLS®) skills curriculum, which was adapted into the urology-specific Basic Laparoscopic Urologic Surgery (BLUS®) skills curriculum by Sweet et al.⁴⁴. A similar curriculum has been designed for robot-assisted surgery, entitled Fundamentals of Robotic Surgery (FRS®)⁴⁵, and is currently undergoing validation.

Furthermore, a number of procedure-specific curricula have also been described and validated. In robotic surgery, the European Association of Urology (EAU) Section of Robotics

(ERUS) have developed a 12-week training curriculum for RARP³⁵. It includes e-learning, one week of structured multi-modality simulation training and supervised modular training. Eight training surgeons took part in the programme and the authors demonstrated face validity, feasibility, acceptability and educational impact. In endourology, the curricula employed by Soria et al.¹⁶ and Brunckhorst et al.¹⁴ constitute important examples.

Team training is an important concept, which has been fairly neglected in the literature. The concepts of full immersion simulation and using high-fidelity simulated ORs can be important means to provide both technical and non-technical skills training for all members of the surgical team. The former modality is especially important, as it is significantly cheaper than the latter. Team training is especially important in robot-assisted surgery as the surgeon is at the console away from the patient and, thus, relies on assistants for the safety of the patient. The Xperience® Team Trainer (XTT; Mimic Technologies) is developed to train both the surgeon and the assistant. Although it is currently used alongside generic skills modules, it is hoped that procedure-specific modules will also be developed. The platform has demonstrated face, content, construct and concurrent validity⁴⁶.

A significant boundary for adoption of simulation training is cost effectiveness of the platforms. A significant proportion of the listed tools are commercially available, save a select few, which are experimental models, developed by institutions. Hence, they are usually high-fidelity models and a significant cost attached to each. The number of validated low-fidelity and inexpensive models are limited. However, where affordable, institutions are strongly recommended to invest in models with evidence-base, as presented.

This article, as with all reviews, has a few limitations. Firstly, despite every effort, relevant studies may have been missed. Although RCTs demonstrate the highest LoE, only a handful of these were identified in the literature. There are no Level 1a studies and the conducted RCTs are all small-scale Level 1b studies. Furthermore, earlier simulators have higher LoE and LoR as more studies have been conducted using them. In contrast, newly developed simulators have not had as thorough evaluation and hold lower LoR. Consequently, it is important that greater emphasis is made on validating the more recently developed simulators.

Very few studies compared the available models to assess strengths and weaknesses of each model. Most studies demonstrate face and content validity, which are subjective measures of validation as opposed to more objective measures including construct, concurrent and predictive validity. Moreover, there is no clear consensus on the exact definitions of validity terms. In the included studies, the use of these terms varied and, therefore, were judged based on the definitions in Figure 2. Another limitation on quality of studies was that majority of studies recruited medical students, possibly due to availability.

Standardisation of validation is an important factor for future studies. It is strongly recommended that authors conform to the definitions of validity terms by McDougall⁵ and van Nortwick et al.⁶ (Figure 2). As per these definitions, content validity should be limited to experts only. Furthermore, studies should mainly recruit residents in-training, as they will ultimately be the first to receive such training. Furthermore, with urology having produced an outstanding number of validation studies, these should be utilised for power calculations prior to studies. Expert(s) are recommended to pilot and explore each tool and a Delphi

technique should be utilised to decide which tasks should be investigated and utilised prior to investigation. The concept of “distributed” training, with a maximum of two sessions performed per day, at least one hour apart has also been recommended in the literature⁴⁷. Finally, authors should discuss the value of their study in light of the modified educational OCEBM classification system for LoE and LoR, as adapted by the European Association of Endoscopic Surgery⁷.

CONCLUSION

With the various changes in surgical training, simulators have become of increasing interest in all surgical specialties. It is of utmost importance to identify which of these models would be most valuable to be implemented in a curriculum for postgraduate training programs. Across the field of urology, there has been a significant increase in the number of simulators developed. However, there are still too few a number of validation studies with high level of evidence. Nevertheless, efforts should be made to utilise the currently available models in a curricular approach.

ACKNOWLEDGEMENTS

PDG and KA acknowledge support from the NIHR Biomedical Research Centre, MRC Centre for Transplantation, King's Health Partners, Guy's and St. Thomas' Charity, School of Surgery, London Deanery, Royal College of Surgeons of England, Intuitive Surgical, The Urology Foundation, Olympus, EU-FP7, Prostate Cancer UK, Technology Strategy Board and The Vattikuti Foundation.

Conflict of Interest

None declared

REFERENCES

1. Halsted SW: The Training Of The Surgeon. Bull John Hopkins Hosp 1904; **15**: 267.
2. Reznick RK, MacRae H. Teaching surgical skills--changes in the wind. N Engl J Med 2006;**355**(25):2664-9.
3. Schout BM, Hendriks AJ, Scherpbier AJ, et al. Update on training models in endourology: a qualitative systematic review of the literature between January 1980 and April 2008. European urology 2008;**54**(6):1247-61.
4. Pan N, Leoncini E, de Belvis G, et al. Evaluation of the endorsement of the preferred reporting items for systematic reviews and meta-analysis (PRISMA) statement on the quality of published systematic review and meta-analyses. PloS one 2013;**8**(12):e83138.
5. McDougall EM. Validation of surgical simulators. J Endourol 2007;**21**(3):244-7.
6. Van Nortwick SS, Lendvay TS, Jensen AR, et al. Methodologies for establishing validity in surgical simulation studies. Surgery 2010;**147**(5):622-30.
7. Carter FJ, Schijven MP, Aggarwal R, et al. Consensus guidelines for validation of virtual reality surgical simulators. Surgical endoscopy 2005;**19**(12):1523-32.
8. Gettman MT, Le CQ, Rangel LJ, et al. Development of a standardized curriculum for teaching cystoscopic skills using a computer-based endourologic simulator. Simulation in healthcare : journal of the Society for Simulation in Healthcare 2009;**4**(2):92-7.
9. Schout BM, Muijtens AM, Hendriks AJ, et al. Acquisition of flexible cystoscopy skills on a virtual reality simulator by experts and novices. BJU Int 2010;**105**(2):234-9.
10. Schout BM, Ananias HJ, Bemelmans BL, et al. Transfer of cysto-urethroscopy skills from a virtual-reality simulator to the operating room: a randomized controlled trial. BJU international 2010;**106**(2):226-31; discussion 31.
11. Bowling CB, Greer WJ, Bryant SA, et al. Testing and validation of a low-cost cystoscopy teaching model: a randomized controlled trial. Obstetrics and gynecology 2010;**116**(1):85-91.
12. Ahmed K, Aydin A, Dasgupta P, et al. A Novel Cadaveric Simulation Program in Urology. J Surg Educ 2015; **72**: 556.
13. Matsumoto ED, Pace KT, D'A Honey RJ. Virtual reality ureteroscopy simulator as a valid tool for assessing endourological skills. International journal of urology : official journal of the Japanese Urological Association 2006;**13**(7):896-901.
14. Brunckhorst O, Shahid S, Aydin A, et al. Simulation-based ureteroscopy skills training curriculum with integration of technical and non-technical skills: a randomised controlled trial. Surg Endosc 2015;**29**(9):2728-35.
15. Matsumoto ED, Hamstra SJ, Radomski SB, et al. The effect of bench model fidelity on endourological skills: a randomized controlled study. The Journal of urology 2002;**167**(3):1243-7.
16. Soria F, Morcillo E, Sanz JL, et al. Description and validation of realistic and structured endourology training model. American journal of clinical and experimental urology 2014;**2**(3):258-65.
17. Mishra S, Kurien A, Ganpule A, et al. Percutaneous renal access training: content validation comparison between a live porcine and a virtual reality (VR) simulation model. BJU international 2010;**106**(11):1753-6.
18. Zhang Y, Ou TW, Jia JG, et al. Novel biologic model for percutaneous renal surgery learning and training in the laboratory. Urology 2008;**72**(3):513-6.
19. Hammond L, Ketchum J, Schwartz BF. A new approach to urology training: a laboratory model for percutaneous nephrolithotomy. J Urol 2004;**172**(5 Pt 1):1950-2.
20. Ballaro A, Briggs T, Garcia-Montes F, et al. A computer generated interactive transurethral prostatic resection simulator. J Urol 1999;**162**(5):1633-5.

21. Schout BM, Bemelmans BL, Martens EJ, et al. How useful and realistic is the uro trainer for training transurethral prostate and bladder tumor resection procedures? *J Urol* 2009;**181**(3):1297-303; discussion 303.
22. Shamim Khan M, Ahmed K, Gavazzi A, et al. Development and implementation of centralized simulation training: evaluation of feasibility, acceptability and construct validity. *BJU Int* 2013;**111**(3):518-23.
23. de Vries AH, van Genugten HG, Hendriks AJ, et al. The Simbla TURBT Simulator in Urological Residency Training: From Needs Analysis to Validation. *J Endourol* 2016; Epub ahead of print.
24. Angulo JC, Arance I, García-Tello A, et al. Virtual reality simulator for training on photoselective vaporization of the prostate with 980 nm diode laser and learning curve of the technique. *Actas urológicas españolas* 2014;**38**(7):451-8.
25. Wijn RP, Persoon MC, Schout BM, et al. Virtual reality laparoscopic nephrectomy simulator is lacking in construct validity. *J Endourol* 2010;**24**(1):117-22.
26. Poniatowski LH, Wolf JS, Jr., Nakada SY, et al. Validity and acceptability of a high-fidelity physical simulation model for training of laparoscopic pyeloplasty. *Journal of endourology / Endourological Society* 2014;**28**(4):393-8.
27. Jiang C, Liu M, Chen J, et al. Construct validity of the chicken crop model in the simulation of laparoscopic pyeloplasty. *Journal of endourology / Endourological Society* 2013;**27**(8):1032-6.
28. Tunitsky E, Murphy A, Barber MD, et al. Development and validation of a ureteral anastomosis simulation model for surgical training. *Female pelvic medicine & reconstructive surgery* 2013;**19**(6):346-51.
29. Sabbagh R, Chatterjee S, Chawla A, et al. Transfer of laparoscopic radical prostatectomy skills from bench model to animal model: a prospective, single-blind, randomized, controlled study. *J Urol* 2012;**187**(5):1861-6.
30. Hung AJ, Ng CK, Patil MB, et al. Validation of a novel robotic-assisted partial nephrectomy surgical training model. *BJU Int* 2012;**110**(6):870-4.
31. Chowriappa A, Raza SJ, Fazili A, et al. Augmented-reality-based skills training for robot-assisted urethrovesical anastomosis: a multi-institutional randomised controlled trial. *BJU Int* 2015;**115**(2):336-45.
32. Kang SG, Cho S, Kang SH, et al. The Tube 3 module designed for practicing vesicourethral anastomosis in a virtual reality robotic simulator: determination of face, content, and construct validity. *Urology* 2014;**84**(2):345-50.
33. Kim JY, Kim SB, Pyun JH, et al. Concurrent and predictive validation of robotic simulator Tube 3 module. *Korean journal of urology* 2015;**56**(11):756-61.
34. Alemozafer M, Narayanan R, Percy AA, et al. Validation of a novel, tissue-based simulator for robot-assisted radical prostatectomy. *J Endourol* 2014;**28**(8):995-1000.
35. Volpe A, Ahmed K, Dasgupta P, et al. Pilot Validation Study of the European Association of Urology Robotic Training Curriculum. *European urology* 2015; **68**: 292.
36. Cabello R, González C, Quicios C, et al. An Experimental Model for Training in Renal Transplantation Surgery With Human Cadavers Preserved Using W. Thiel's Embalming Technique. *J Surg Educ* 2015; **72**: 192.
37. Brunckhorst O, Shahid S, Aydin A, et al. The Relationship Between Technical And Nontechnical Skills Within A Simulation-Based Ureteroscopy Training Environment. *Journal of surgical education* 2015;**72**(5):1039-44.
38. Brewin J, Tang J, Dasgupta P, et al. Full immersion simulation: validation of a distributed simulation environment for technical and non-technical skills training in Urology. *BJU Int* 2015;**116**(1):156-62.
39. Lee JY, Mucksavage P, Canales C, et al. High fidelity simulation based team training in urology: a preliminary interdisciplinary study of technical and nontechnical skills in laparoscopic complications management. *J Urol* 2012;**187**(4):1385-91.

40. Abdelshahid CS, Quach S, Nelson C, et al. High-fidelity simulation-based team training in urology: evaluation of technical and nontechnical skills of urology residents during laparoscopic partial nephrectomy. *Journal of surgical education* 2013;**70**(5):588-95.
41. Abboudi H, Khan MS, Aboumarzouk O, et al. Current status of validation for robotic surgery simulators - a systematic review. *BJU Int* 2013;**111**(2):194-205.
42. Persoon MC, van Putten K, Muijtjens AM, et al. Effect of distraction on the performance of endourological tasks: a randomized controlled trial. *BJU international* 2011;**107**(10):1653-7.
43. Hung AJ, Patil MB, Zehnder P, et al. Concurrent and predictive validation of a novel robotic surgery simulator: a prospective, randomized study. *J Urol* 2012;**187**(2):630-7.
44. Sweet RM, Beach R, Sainfort F, et al. Introduction and validation of the American Urological Association Basic Laparoscopic Urologic Surgery skills curriculum. *J Endourol* 2012;**26**(2):190-6.
45. Smith R, Patel V, Satava R. Fundamentals of robotic surgery: a course of basic robotic surgery skills based upon a 14-society consensus template of outcomes measures and curriculum development. *Int J Med Robot* 2014;**10**(3):379-84.
46. Xu S, Perez M, Perrenot C, et al. Face, content, construct, and concurrent validity of a novel robotic surgery patient-side simulator: the Xperience™ Team Trainer. *Surg Endosc* 2015; Epub ahead of print.
47. Mackay S, Morgan P, Datta V, et al. Practice distribution in procedural skills training: a randomized controlled trial. *Surg Endosc* 2002;**16**(6):957-61.

FIGURE LEGENDS

Figure 1: Study selection process, according to the Preferred Reporting Items for Systematic reviews And meta-analysis (PRISMA) statement

Figure 2: Definitions of Validity, based on definitions by McDougall et al.⁵ and van Nortwick et al.⁶.

Abbreviation: OSATS – Objective Structured Assessment of Technical Skills

TABLES**Table 1:** Validation studies on urological training models (1990–2015) for Urethrocystoscopy (UCS), Ureterorenoscopy (URS) and Percutaneous access and/or nephrolithotomy (PCNL). Abbreviations: LoE- Level of Evidence, LoR- Level of Recommendation, VR- virtual reality.

*models also used to perform intravesical botulinum toxin injections, •model also used for bladder biopsy

Procedure	Name of Model	Institution / Manufacturer	Type of Model	Study	Validation	Participants		LoE	LoR
						n	Demographics		
UCS	ETXY-Uro Adam	ProDelphus, Olinda – PE, Brazil	Bench	Soria et al. (2014)	Face, Content, Construct B	50	40 Trainees, 10 Experts	2b	3
	URO-Trainer	Karl Storz GmbH, Tuttlingen, Germany	VR	Reich et al. (2006)	Content, Construct A, Construct B	36	24 Medical Students, 12 Residents	2b	3
	URO Mentor	Simbionix, Lod, Israel	VR	Wilhelm et al. (2002)	Construct A, Construct B	21	Medical Students	2a	1
				Shah et al. (2002)	Construct A	14	Urology Nurses	2c	
				Shah et al. (2002)	Construct A, Construct B	17	10 Novices, 7 Experts	2b	
				Gettman et al. (2008)	Face, Construct B	57	30 Novices, 27 Experts	2b	
				Gettman et al. (2009)	Face, Construct A	10	Novices	2c	
				Dolmans et al. (2009)	Face, Content	89	33 Referent, 56 Experts	4	
				Schout et al. (2010a)	Construct B	100	100 Interns	1b	
				Schout et al. (2010b)	Construct B	80	50 Novices, 30 Experts	2b	
				Persoon et al. (2011)	Construct A	86	Medical Students	1b	
				Shamim Khan et al. (2013)	Face, Content, Construct B	33	33 Trainees	2b	
				Zhang et al. (2013)	Construct A	18	Urologists	2c	
	URO Mentor + Glass globe	Catharina Hospital Eindhoven, Eindhoven, Netherlands	VR + Bench	Persoon et al. (2010)	Face, Construct B	32	Medical Students	2a	3
	URO Mentor + Uro Scopic Trainer	Simbionix, Lod, Israel + Limbs and Things, Bristol, UK	VR + Bench	Matsumoto et al. (2006a)	Face, Construct B	16	Residents	2c	3

URS	Boar Urinary tract	Department of Urology, Mayo Clinic, Arizona, Phoenix, AZ	Animal●	Grimsby et al. (2011)	Construct A	2	Residents	2c	4
	Live Porcine	Minimally Invasive Surgery Centre Jesús Usón Cáceres, Spain	Animal	Soria et al. (2014)	Face, Content	50	40 Trainees, 10 Experts	2b	3
	Fresh Frozen Cadaver	University of Alabama at Birmingham, Birmingham, Alabama	Cadaver	Bowling et al. (2010)	Construct B	29	Obstetric Residents	1b	2
		British Association of Urological Surgeons, UK	Cadaver●*	Ahmed et al. (2015)	Face, Content	102	75 Participants (Trainees, Specialists), 27 Experts	4	
	Penrose drain	University of Toronto, Canada	Bench	Matsumoto et al. (2002)	Construct A	40	Medical Students	1b	2
	Scope Trainer	Mediskills, Northampton, UK	Bench	Brehmer et al. (2002)	Face, Content, Construct B	14	5 Trainees, 9 Consultants	2b	3
				Brehmer et al. (2005)	Construct A	26	26 Residents	2c	
	Adult Ureteroscopy Trainer	Ideal Anatomic Modelling, Holt, Michigan, USA	Bench	White et al. (2010)	Face, Content, Construct B	46	Medical Students, Residents, Representatives	2b	3
	Endo-Urologie Modell	Karl Storz GmbH, Tuttlingen, Germany	Bench	Mishra et al. (2011)	Face, Construct A	21	Urologists	2b	3
	Cook URS Model	Cook Medical, Bloomington, IN, USA	Bench	Blankstein et al. (2015)	Face, Content, Construct A	15	Junior Residents	2c	3
	Uro Scopic Trainer	Limbs and Things, Bristol, UK	Bench	Matsumoto et al. (2002)	Construct A	40	Medical Students	1b	2
				Matsumoto et al. (2001)	Construct A, Construct B	17	Residents	2c	
				Mishra et al. (2011)	Face, Construct A	21	Urologists	2b	
			Bench + Porcine	Chou et al. (2006)	Concurrent	16	Medical Students	2a	
	ETXY-Uro Adam + Porcine renoureteral tissue + Live Porcine URO Mentor	Minimally Invasive Surgery Centre Jesús Usón Cáceres, Spain	Bench + Animal	Soria et al. (2014)	Face, Content, Construct A, Construct B	50	40 Trainees, 10 Experts	2b	3
		Simbionix, Lod, Israel	VR	Wilhelm et al. (2002)	Construct A, Construct B	21	Medical Students	2a	2

PCNL	URO Mentor + Uro Scopic Trainer	Simbionix, Lod, Israel + Limbs & Things Ltd, Bristol, UK	VR + Bench	Michel et al. (2002)	Face, Content	-	Urologists	4	2
				Watterson et al. (2002)	Face, Construct A	20	Novices	2a	
				Jacomides et al. (2004)	Construct A, Construct B	32	16 Students, 16 Residents	2b	
				Knoll et al. (2005)	Construct B	20	Urologists	2c	
				Dolmans et al. (2009)	Face, Content	52	20 Referents, 32 Experts		
				Mishra et al. (2011)	Face, Construct A	21	Urologists	2b	
				Cai et al. (2013)	Construct A	30	17 Attending physicians, 13 Associate Chief Physicians	2c	
				Chou et al. (2006)	Concurrent	16	Medical Students	2a	
				Ogan et al. (2004)	Construct B, Predictive	32	16 Students, 16 residents	2b	
				Matsumoto et al. (2006a)	Predictive	16	16 Medical Students	2c	
	Fresh Frozen Cadavers	British Association of Urological Surgeons, UK	Cadaver	Matsumoto et al. (2006b)	Face, Construct B	16	Residents	2b	N/A
				Brunckhorst et al. (2015)	Face, Construct B	32	Medical Students	2a	
				Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4	
				Veneziano et al. (2015)	Face, Content	14	Urologists	4	
	C-arm Trainer (CAT)	SimPORTAL, University of Minnesota, MN, USA	Bench	Knudsen et al. (2006)	Construct A	63	31 Medical Students, 31 Residents, 1 Fellow	2a	2
				Papatsoris et al. (2012)	Construct A	36	Trainees	2c	
				Shamim Khan et al. (2013)	Face, Content, Construct B	33	33 Trainees	2b	
				Mishra et al. (2010a)	Face, Content, Construct B, Predictive	24	15 Novices, 9 Experts	2b	
				Mishra et al. (2010b)	Content	24	Experts	4	
	PERC Mentor	Simbionix, Lod, Israel	VR						

Porcine kidney model	Xuanwu Hospital of Capital Medical University and Peking University, China	Animal	Zhang et al. (2008)	Face	42	Urologists	4	N/A
Porcine kidney in chicken carcass model	Southern Illinois University School of Medicine, Springfield, IL, USA	Animal	Hammond et al. (2004)	Face	-	Residents	4	N/A

Table 2: Validation studies on urological training models (1990–2015) for Transurethral resection and laser prostate therapies. Abbreviations: LoE- Level of Evidence, LoR- Level of Recommendation, TURP - Transurethral resection of the prostate, VR- virtual reality, TURBT - Transurethral resection of bladder tumours, HoLEP - Holmium laser enucleation of the prostate, PVP - Photoselective vaporisation of the prostate, VR- virtual reality.

Procedure	Model	Institution / Manufacturer	Type of Model	Study	Validation	Participants		LoE	LoR
						n	Demographics		
TURP	Bristol TURP Trainer	Limbs & Things Ltd, Bristol, UK	Bench	Shamim Khan et al. (2013)	Face, Content, Construct B	33	33 Trainees	2b	2
				Brewin et al. (2014)	Face, Content, Construct B	16	8 Novices, 8 Experts	2b	
				Brewin et al. (2015)	Face, Content, Construct B	20	10 Trainees, 10 Experts	2b	
	SurgicalSIM TURP (UW TURP Trainer)	METI, Seattle, WA, USA	VR	Sweet et al. (2004)	Face, Content, Construct B	91	19 Novices, 72 Certified Urologists	2b	2
				Rashid et al. (2007)	Construct B	13	72 Urologists, 45 Residents, 19 Novices	2b	
				Hudak et al. (2010)	Construct B	35	4 Medical Students, 19 Trainees, 12 Residents	2b	
	TURP Simulator UroSim/TURPSim	University College London, UK VirtaMed AG, Zurich, Switzerland	VR	Ballaro et al. (1999)	Content	3	3 Urologist	4	N/A
			VR	Bright et al. (2012)	Content, Construct B	18	11 Novices, 7 Experts	2b	2
				Zhu et al. (2013)	Face, Construct A	38	Urology Trainees	2a	
	PelvicVision	Melerit Medical AB, Linkoping, Sweden	VR	Kishore et al. (2009)	Construct A	18	Medical Students	2a	
				Kallstrom et al. (2005)	Content, Construct A	44	37 Urologist, 7 Students	2c	3
TURBT	Uro Trainer	Karl Storz GmbH, Tuttlingen, Germany	VR	Kallstrom et al. (2010a)	Construct A	24	Urologists	2c	
				Kallstrom et al. (2010b)	Face, Content, Construct B	20	11 Medical Students, 9 Urologists	2b	
				Schout et al. (2009)	Face, Content	97	20 Experts, 77 Novices	4	4
	Fresh Frozen Cadavers	British Association of Urological Surgeons, UK	Cadaver	Mishra et al. (2010c)	Face, Content	19	10 Experts, 9 Novices	4	
				Ahmed et al. (2015)	Face, Content	10	75 Participants, 27 Experts	4	N/A
						2			
	Bristol TURBT Model	Limbs & Things Ltd, Bristol, UK	Bench	Shamim Khan et al. (2013)	Face, Content, Construct B	33	33 Trainees	2b	3
				de Vries et al. (2015)	Face, Content, Construct B	76	25 Novices, 26 Intermediates, 25	2b	3
	Simbla TURBT Simulator	SAMED GmbH, Dresden, Germany	Bench						

	URO-Trainer	Karl Storz GmbH, Tuttlingen, Germany	VR	Reich et al. (2006)	Content, Construct A, Construct B	36	Experts 24 Medical Students, 12 Residents	2b	3
				Schout et al. (2009)	Face, Content	64	19 Experts, 45 Novices	4	
				Kruck et al. (2011)	Construct A	15	Residents		
HoLEP	Holmium Surgical Simulator	Kansai Medical University, Japan	Bench	Aydin et al. (2014)	Face, Content	36	13 Urology Trainees, 23 Senior Urologists	4	N/A
	UroSim	VirtaMed AG, Zurich, Switzerland	VR	Kuronen-Stewart et al. (2015)	Face, Content, Construct B	42	6 Experts, 18 Trainees, 18 Novices	2b	3
PVP	GreenLight Simulator	American Medical Systems Inc, Minnetonka, MN, USA	VR	Herlemann et al. (2013)	Face, Construct B	18	9 Novices, 9 Advanced users	2b	2
				Aydin et al. (2015)	Face, Content, Construct A, Construct B	46	25 Novices, 14 Intermediates, 7 Experts	2b	
Diode PVP	MyoSim	VirtMed AG, Zurich, Switzerland	VR	Noureldin et al. (2014)	Construct A, Construct B	25	25 Trainees	2b	
				Angulo et al. (2014)	Construct B	18	Medical students, residents, specialists	2b	3

Table 3: Validation studies on urological training models (1990–2015) for Laparoscopic Urology. Abbreviations: UVA- urethro-vesical anastomosis, LoE- Level of Evidence, LoR- Level of Recommendation, VR- virtual reality.

Procedure	Model	Institution / Manufacturer	Type of Model	Study	Validation	Participants		LoE	LoR
						n	Demographics		
Partial/Radical Nephrectomy	Procedicus MIST Nephrectomy Simulator	Mentice, Gothenburg, Sweden	VR	Brewin et al. (2010)	Face, Content, Construct B	28	8 Experts, 10 Novices, 10 Trainees	2b	2
				Shamim Khan et al. (2013)	Face, Content, Construct B	33	33 Trainees	2b	
				Wijn et al. (2010)	Construct B	64	22 Novices, 32 Intermediates, 10 Experienced	2b	
	Partial Nephrectomy Renal Model	University of California Irvine Medical Center, California, USA	Bench	Fernandez et al. (2012)	Face, Content	5	5 Urology Fellows	4	2
				Lee et al. (2012)	Construct B	8	8 Residents	2b	
				Abdelshehid et al. (2013)	Construct B	18	9 Residents	2b	
	Rabbit model	University Hospital Gasthuisberg, Belgium	Animal	Molinas et al. (2004)	Construct B	20	10 Medical Students, 10 Gynaecologists	2b	3
	Porcine Kidney		Animal	De Win et al. (2013)	Content, Construct A, Predictive	22	Medical Students	2b	3
Pyeloplasty	Pyeloplasty Simulator Model	University of Minnesota, Minnesota, USA	Bench	Poniatowski et al. (2014)	Face, Content, Construct B	31	Clinical Urologists	2b	3
	3D Silicon Model	The Hospital for Sick Children, Toronto, Ontario, Canada	Bench	Cheung et al. (2014)	Face, Content	27	24 Paediatric Urology Fellows, 3 Faculty Members	4	N/A
	Porcine Bladder	Department of Urology, SLK Klinikum Heilbronn, Germany	Animal	Teber et al. (2010)	Construct A	5	5 Laparoscopic Surgeons	3	4
	Chicken crop model	Sun Yat-sen Memorial Hospital, Sun Yat-sen University, Guangzhou, Guangdong, China	Animal	Jiang et al. (2013)	Construct B	15	5 Experts, 5 Specialists, 5 Junior Residents	2b	3
Ureteral Re-implantation	Hydrogel model	Cleveland Clinic, Cleveland, OH, USA	Bench	Tunitsky et al. (2013)	Face, Content, Construct B	20	12 Trainees, 5 Robotic Experts, 4 Procedure Experts	2b	3

UVA	Latex UV Model (UVM)	McMaster University, Hamilton, Ontario, Canada	Bench	Sabbagh et al. (2012)	Face, Predictive	28	Senior Residents, Fellows, Staff Surgeons	2a	3
	Live Porcine	McMaster University, Hamilton, Ontario, Canada	Animal	Sabbagh et al. (2012)	Face	28	Senior Residents, Fellows, Staff Surgeons	2a	3
	Porcine Intestine	Baylor College of Medicine, Houston, Texas, USA	Animal	Boon et al. (2008)	Construct B	12	Residents, Medical Students	2b	3
	Chicken Chest Model	AMC University Hospital, Amsterdam, Netherlands	Animal	Laguna et al. (2006)	Construct B	5	Urologists	2c	4

Table 4: Validation studies on urological training models (1990–2015) for Robot-assisted Urological Surgery. Abbreviations: LoE- Level of Evidence, LoR- Level of Recommendation, RAPN- robot-assisted partial nephrectomy, UVA- urethro-vesical anastomosis, RARP- robot-assisted radical prostatectomy, VR- virtual reality, HoST- Hands-on Surgical Training, RoSS- Robotic Surgical Simulator, GU- genitourinary, ERUS- EAU Robotic Urology section.

Procedure	Model	Institution / Manufacturer	Type of Model	Study	Validation	Participants		LoE	LoR
						n	Demographics		
RAPN	Porcine kidney + Styrofoam ball	USC Institute of Urology, University of Southern California, Los Angeles, CA, USA	Animal	Hung et al. (2012a)	Face, Content, Construct B	46	24 Novices, 9 Intermediates, 13 Experts	2b	2
			+ Cadaver	Hung et al. (2012b)	Concurrent, Predictive	24	2 Students, 1 Urology Intern, 14 Residents, 5 Fellows, 2 Urology staff	1b	
	dV-Trainer	Mimic Technologies Inc., Seattle, WA, USA	AR + VR + Animal	Hung et al. (2015)	Face, Content, Construct B, Concurrent	42	15 novices, 13 intermediates, 14 experts	2b	3
UVA	HoST/RoSS	Simulated Surgical Systems; Williamsville, NY, USA	AR + VR	Chowriappa et al. (2015)	Face, Concurrent	52	30 Fellows, 22 Residents	1b	3
	Tube 3/dV-Trainer	Mimic Technologies, Inc., Seattle, WA, USA	VR	Kang et al. (2014)	Face, Content, Construct B	20	10 Novices, 10 Experienced Surgeons	2b	3
				Kim et al. (2015)	Concurrent, Predictive	11	8 Residents, 3 Fellows	2c	
RARP	Porcine GU Model	University of Southern California, Los Angeles, CA, USA	Animal	Alemozaffar et al. (2014)	Face, Content, Construct B	20	10 Novices, 10 Experts	2c	4
	ERUS Curriculum	European Association of Urology	VR + Bench + Animal + Cadaver	Volpe et al. (2014)	Face	12	10 Experts, 2 Fellows	2c	4

Table 5: Validation studies of urological training models for Open Urological Surgery (1990–2015). Abbreviations: LoE- Level of Evidence, SPC- Suprapubic Catheterisation.

Procedure	Model	Institution / Manufacturer	Type of Model	Study	Validation	Participants		LoE
						n	Demographics	
Suprapubic Catheterisation	SPC Training model	Centre for Education in Medicine, Northwestern University Feinberg School of Medicine, Chicago, IL	Bench	Singal et al. (2015)	Face, Content	25	25 General Surgeons	4
	SPC Training model	Western Hospital, Melbourne, Australia	Bench	Hossack et al. (2013)	Face	30	30 Trainees	4
	UroEmerge SPC model	St Bartholomew's and The Royal London Hospitals, London, UK	Bench	Shergill et al. (2008)	Construct A, Predictive	36	36 Junior Trainees	2c
Circumcision	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants (Trainees, Specialists), 27 Experts	4
Vasectomy	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Vasovasostomy	Silicone tube + Rat vas deferens	University of Toronto, Ontario, Canada	Bench + Animal	Grober et al. (2004)	Construct A, Predictive	13	Residents	2c
Testicular Fixation	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Hydrocele Repair	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Radical Orchiectomy	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Open Cystostomy	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Management of Bladder Perforation	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Transureteroureterostomy	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4

Boari Flap / Psoas Hitch	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Open Surgical Packing of Pelvis	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Nephrectomy	Fresh Frozen Cadaver	British Association of Urological Surgeons, UK	Cadaver	Ahmed et al. (2015)	Face, Content	102	75 Participants, 27 Experts	4
Renal Graft Transplant	Thiel-Embalmed Cadaver	Universidad Autonoma de Madrid, Madrid, Spain	Cadaver	Cabello et al. (2014)	Face	28	Residents, Junior Transplant Surgeons	4

Face Validity – Opinions, including of non-experts, regarding the realism of the simulator

Content Validity – Opinions of experts about the simulator and its appropriateness for training

Construct Validity

- **A** – one group – Ability of the simulator to assess and differentiate between the level of experience of an individual or group measured over time
- **B** – between groups – Ability of the simulator to distinguish between different levels of experience

Concurrent Validity – Comparison of the new model against the older and gold standard, usually by OSATS

Predictive Validity – Correlation of performance with operating room performance, usually measured by OSATS

LoE – Level of Evidence

LoR – Level of Recommendation

RCT – Randomised Controlled Trial

VR – Virtual Reality

FFC – Fresh Frozen Cadaver

TURP – transurethral resection of the prostate

TURBT – transurethral resection of bladder tumours

UVA – urethro-vesical anastomosis

RARP – robot-assisted radical prostatectomy

ALL INCLUDED STUDIES

- ABDELSHEHID, C. S., QUACH, S., NELSON, C., GRAVERSEN, J., LUSCH, A., ZARRAGA, J., ALIPANAH, R., LANDMAN, J. & MCDUGALL, E. M. 2013. High-fidelity simulation-based team training in urology: evaluation of technical and nontechnical skills of urology residents during laparoscopic partial nephrectomy. *J Surg Educ*, 70, 588-95.
- AHMED, K., AYDIN, A., DASGUPTA, P., KHAN, M. S. & MCCABE, J. E. 2015. A Novel Cadaveric Simulation Program in Urology. *J Surg Educ*.
- ALEMOZAFFAR, M., NARAYANAN, R., PERCY, A. A., MINNILLO, B. B., STEINBERG, P., HALEBLIAN, G., GAUTAM, S., MATTHES, K. & WAGNER, A. A. 2014. Validation of a novel, tissue-based simulator for robot-assisted radical prostatectomy. *J Endourol*, 28, 995-1000.
- ANGULO, J. C., ARANCE, I., GARCIA-TELLO, A., LAS HERAS, M. M., ANDRES, G., GIMBERNAT, H., LISTA, F. & RAMON DE FATA, F. 2014. Virtual reality simulator for training on photoselective vaporization of the prostate with 980 nm diode laser and learning curve of the technique. *Actas Urol Esp*, 38, 451-8.
- AYDIN, A., AHMED, K., BREWIN, J., KHAN, M. S., DASGUPTA, P. & AHO, T. 2014. Face and content validation of the prostatic hyperplasia model and holmium laser surgery simulator. *J Surg Educ*, 71, 339-44.
- AYDIN, A., MUIR, G. H., GRAZIANO, M. E., KHAN, M. S., DASGUPTA, P. & AHMED, K. 2015. Validation of the GreenLight Simulator and development of a training curriculum for photoselective vaporisation of the prostate. *BJU Int*, 115, 994-1003.
- BALLARO, A., BRIGGS, T., GARCIA-MONTES, F., MACDONALD, D., EMBERTON, M. & MUNDY, A. R. 1999. A computer generated interactive transurethral prostatic resection simulator. *J Urol*, 162, 1633-5.
- BLANKSTEIN, U., LANTZ, A. G., RJ, D. A. H., PACE, K. T., ORDON, M. & LEE, J. Y. 2015. Simulation-based flexible ureteroscopy training using a novel ureteroscopy part-task trainer. *Can Urol Assoc J*, 9, 331-5.
- BOON, J. R., SALAS, N., AVILA, D., BOONE, T. B., LIPSHULTZ, L. I. & LINK, R. E. 2008. Construct validity of the pig intestine model in the simulation of laparoscopic urethrovesical anastomosis: tools for objective evaluation. *J Endourol*, 22, 2713-6.
- BOWLING, C. B., GREER, W. J., BRYANT, S. A., GLEASON, J. L., SZYCHOWSKI, J. M., VARNER, R. E., HOLLEY, R. L. & RICHTER, H. E. 2010. Testing and validation of a low-cost cystoscopy teaching model: a randomized controlled trial. *Obstet Gynecol*, 116, 85-91.
- BREHMER, M. & SWARTZ, R. 2005. Training on bench models improves dexterity in ureteroscopy. *Eur Urol*, 48, 458-63; discussion 463.
- BREHMER, M. & TOLLEY, D. 2002. Validation of a bench model for endoscopic surgery in the upper urinary tract. *Eur Urol*, 42, 175-9; discussion 180.
- BREWIN, J., AHMED, K., KHAN, M. S., JAYE, P. & DASGUPTA, P. 2014. Face, content, and construct validation of the Bristol TURP trainer. *J Surg Educ*, 71, 500-5.
- BREWIN, J., NEDAS, T., CHALLACOMBE, B., ELHAGE, O., KEISU, J. & DASGUPTA, P. 2010. Face, content and construct validation of the first virtual reality laparoscopic nephrectomy simulator. *BJU Int*, 106, 850-4.
- BREWIN, J., TANG, J., DASGUPTA, P., KHAN, M. S., AHMED, K., BELLO, F., KNEEBONE, R. & JAYE, P. 2015. Full immersion simulation: validation of a distributed simulation environment for technical and non-technical skills training in Urology. *BJU Int*, 116, 156-62.
- BRIGHT, E., VINE, S., WILSON, M. R., MASTERS, R. S. & MCGRATH, J. S. 2012. Face validity, construct validity and training benefits of a virtual reality TURP simulator. *Int J Surg*, 10, 163-6.
- BRUNCKHORST, O., SHAHID, S., AYDIN, A., MCILHENNY, C., KHAN, S., RAZA, S. J., SAHAI, A., BREWIN, J., BELLO, F., KNEEBONE, R., KHAN, M. S., DASGUPTA, P. & AHMED, K. 2015. Simulation-based ureteroscopy skills training curriculum with integration of technical and non-technical skills: a randomised controlled trial. *Surg Endosc*, 29, 2728-35.
- CABELLO, R., GONZALEZ, C., QUICIOS, C., BUENO, G., GARCIA, J. V., ARRIBAS, A. B. & CLASCA, F. 2014. An Experimental Model for Training in Renal Transplantation Surgery With Human Cadavers Preserved Using W. Thiel's Embalming Technique. *J Surg Educ*.

- CAI, J. L., ZHANG, Y., SUN, G. F., LI, N. C., YUAN, X. L. & NA, Y. Q. 2013. Proficiency of virtual reality simulator training in flexible retrograde ureteroscopy renal stone management. *Chin Med J (Engl)*, 126, 3940-3.
- CHEUNG, C. L., LOOI, T., LENDVAY, T. S., DRAKE, J. M. & FARHAT, W. A. 2014. Use of 3-dimensional printing technology and silicone modeling in surgical simulation: development and face validation in pediatric laparoscopic pyeloplasty. *J Surg Educ*, 71, 762-7.
- CHOU, D. S., ABDELSHEHID, C., CLAYMAN, R. V. & MCDOUGALL, E. M. 2006. Comparison of results of virtual-reality simulator and training model for basic ureteroscopy training. *J Endourol*, 20, 266-71.
- CHOWRIAPPA, A., RAZA, S. J., FAZILI, A., FIELD, E., MALITO, C., SAMARASEKERA, D., SHI, Y., AHMED, K., WILDING, G., KAOUK, J., EUN, D. D., GHAZI, A., PEABODY, J. O., KESAVADAS, T., MOHLER, J. L. & GURU, K. A. 2015. Augmented-reality-based skills training for robot-assisted urethrovesical anastomosis: a multi-institutional randomised controlled trial. *BJU Int*, 115, 336-45.
- DE VRIES, A. H., VAN GENUGTEN, H. G., HENDRIKX, A. J., KOLDEWIJN, E. L., SCHOUT, B. M., TJIAM, I. M., VAN MERRIENBOER, J. J., MUIJTJENS, A. M. & WAGNER, C. 2015. The Simbla TURBT Simulator in Urological Residency Training: From Needs Analysis to Validation. *J Endourol*.
- DE WIN, G., VAN BRUWAENE, S., ALLEN, C. & DE RIDDER, D. 2013. Design and implementation of a proficiency-based, structured endoscopy course for medical students applying for a surgical specialty. *Adv Med Educ Pract*, 4, 103-15.
- DOLMANS, V. E., SCHOUT, B. M., DE BEER, N. A., BEMELMANS, B. L., SCHERPBIER, A. J. & HENDRIKX, A. J. 2009. The virtual reality endourologic simulator is realistic and useful for educational purposes. *J Endourol*, 23, 1175-81.
- FERNANDEZ, A., CHEN, C., MOORE, J., CHEUNG, C., ERDELJAN, P., FULLER, A., MCDOUGALL, E. M., PETERS, T. M. & PAUTLER, S. E. 2012. First Prize: A Phantom Model as a Teaching Modality for Laparoscopic Partial Nephrectomy. *J Endourol*, 26, 1-5.
- GETTMAN, M. T., LE, C. Q., RANGEL, L. J., SLEZAK, J. M., BERGSTALH, E. J. & KRAMBECK, A. E. 2008. Analysis of a computer based simulator as an educational tool for cystoscopy: subjective and objective results. *J Urol*, 179, 267-71.
- GETTMAN, M. T., LE, C. Q., RANGEL, L. J., SLEZAK, J. M., BERGSTALH, E. J. & KRAMBECK, A. E. 2009. Development of a standardized curriculum for teaching cystoscopic skills using a computer-based endourologic simulator. *Simul Healthc*, 4, 92-7.
- GRIMSBY, G. M., ANDREWS, P. E., CASTLE, E. P., WOLTER, C. E., PATEL, B. M. & HUMPHREYS, M. R. 2011. Urologic surgical simulation: an endoscopic bladder model. *Simul Healthc*, 6, 352-5.
- GROBER, E. D., HAMSTRA, S. J., WANZEL, K. R., REZNICK, R. K., MATSUMOTO, E. D., SIDHU, R. S. & JARVI, K. A. 2004. Laboratory based training in urological microsurgery with bench model simulators: a randomized controlled trial evaluating the durability of technical skill. *J Urol*, 172, 378-81.
- HAMMOND, L., KETCHUM, J. & SCHWARTZ, B. F. 2004. A new approach to urology training: a laboratory model for percutaneous nephrolithotomy. *J Urol*, 172, 1950-2.
- HERLEMANN, A., STRITTMATTER, F., BUCHNER, A., KARL, A., REICH, O., BACHMANN, A., STIEF, C. G. & GRATZKE, C. 2013. Virtual reality systems in urologic surgery: an evaluation of the GreenLight simulator. *Eur Urol*, 64, 687-8.
- HOSSACK, T., CHRIS, B. B., BEER, J. & THOMPSON, G. 2013. A cost-effective, easily reproducible, suprapubic catheter insertion simulation training model. *Urology*, 82, 955-8.
- HUDAK, S. J., LANDT, C. L., HERNANDEZ, J. & SODERDAHL, D. W. 2010. External validation of a virtual reality transurethral resection of the prostate simulator. *J Urol*, 184, 2018-22.
- HUNG, A. J., NG, C. K., PATIL, M. B., ZEHNDER, P., HUANG, E., ARON, M., GILL, I. S. & DESAI, M. M. 2012a. Validation of a novel robotic-assisted partial nephrectomy surgical training model. *BJU Int*, 110, 870-4.
- HUNG, A. J., PATIL, M. B., ZEHNDER, P., CAI, J., NG, C. K., ARON, M., GILL, I. S. & DESAI, M. M. 2012b. Concurrent and predictive validation of a novel robotic surgery simulator: a prospective, randomized study. *J Urol*, 187, 630-7.
- HUNG, A. J., SHAH, S. H., DALAG, L., SHIN, D. & GILL, I. S. 2015. Development and Validation of a Novel Robotic Procedure Specific Simulation Platform: Partial Nephrectomy. *J Urol*.
- JACOMIDES, L., OGAN, K., CADEDDU, J. A. & PEARLE, M. S. 2004. Use of a virtual reality simulator for ureteroscopy training. *J Urol*, 171, 320-3; discussion 323.

- JIANG, C., LIU, M., CHEN, J., WANG, P., LIN, T., XU, K., HAN, J., HUANG, H. & HUANG, J. 2013. Construct validity of the chicken crop model in the simulation of laparoscopic pyeloplasty. *J Endourol*, 27, 1032-6.
- KALLSTROM, R., HJERTBERG, H., KJOLHEDE, H. & SVANVIK, J. 2005. Use of a virtual reality, real-time, simulation model for the training of urologists in transurethral resection of the prostate. *Scand J Urol Nephrol*, 39, 313-20.
- KALLSTROM, R., HJERTBERG, H. & SVANVIK, J. 2010a. Construct validity of a full procedure, virtual reality, real-time, simulation model for training in transurethral resection of the prostate. *J Endourol*, 24, 109-15.
- KALLSTROM, R., HJERTBERG, H. & SVANVIK, J. 2010b. Impact of virtual reality-simulated training on urology residents' performance of transurethral resection of the prostate. *J Endourol*, 24, 1521-8.
- KANG, S. G., CHO, S., KANG, S. H., HAIDAR, A. M., SAMAVEDI, S., PALMER, K. J., PATEL, V. R. & CHEON, J. 2014. The Tube 3 module designed for practicing vesicourethral anastomosis in a virtual reality robotic simulator: determination of face, content, and construct validity. *Urology*, 84, 345-50.
- KIM, J. Y., KIM, S. B., PYUN, J. H., KIM, H. K., CHO, S., LEE, J. G., KIM, J. J., CHEON, J., KANG, S. H. & KANG, S. G. 2015. Concurrent and predictive validation of robotic simulator Tube 3 module. *Korean J Urol*, 56, 756-61.
- KISHORE, T. A., BEDDINGFIELD, R., HOLDEN, T., SHEN, Y., REIHSSEN, T. & SWEET, R. M. 2009. Task deconstruction facilitates acquisition of transurethral resection of prostate skills on a virtual reality trainer. *J Endourol*, 23, 665-8.
- KNOLL, T., TROJAN, L., HAECKER, A., ALKEN, P. & MICHEL, M. S. 2005. Validation of computer-based training in ureterorenoscopy. *BJU Int*, 95, 1276-9.
- KNUDSEN, B. E., MATSUMOTO, E. D., CHEW, B. H., JOHNSON, B., MARGULIS, V., CADEDDU, J. A., PEARLE, M. S., PAUTLER, S. E. & DENSTEDT, J. D. 2006. A randomized, controlled, prospective study validating the acquisition of percutaneous renal collecting system access skills using a computer based hybrid virtual reality surgical simulator: phase I. *J Urol*, 176, 2173-8.
- KRUCK, S., BEDKE, J., HENNENLOTTER, J., AMEND, B., MERSEBURGER, A., STENZL, A. & SIEVERT, K. D. 2011. Virtual bladder tumor transurethral resection: an objective evaluation tool to overcome learning curves with and without photodynamic diagnostics. *Urol Int*, 87, 138-42.
- KURONEN-STEWART, C., AHMED, K., AYDIN, A., CYNK, M., MILLER, P., CHALLACOMBE, B., KHAN, M. S., DASGUPTA, P., AHO, T. F. & POPERT, R. 2015. Holmium Laser Enucleation of the Prostate: Simulation-Based Training Curriculum and Validation. *Urology*, 86, 639-46.
- LAGUNA, M. P., ARCE-ALCAZAR, A., MOCHTAR, C. A., VAN VELTHOVEN, R., PELTIER, A. & DE LA ROSETTE, J. J. 2006. Construct validity of the chicken model in the simulation of laparoscopic radical prostatectomy suture. *J Endourol*, 20, 69-73.
- LEE, J. Y., MUCKSAVAGE, P. & MCDUGALL, E. M. 2012. Simulating laparoscopic renal hilar vessel injuries: preliminary evaluation of a novel surgical training model for residents. *J Endourol*, 26, 393-7.
- MATSUMOTO, E. D., HAMSTRA, S. J., RADOMSKI, S. B. & CUSIMANO, M. D. 2001. A novel approach to endourological training: training at the Surgical Skills Center. *J Urol*, 166, 1261-6.
- MATSUMOTO, E. D., HAMSTRA, S. J., RADOMSKI, S. B. & CUSIMANO, M. D. 2002. The effect of bench model fidelity on endourological skills: a randomized controlled study. *J Urol*, 167, 1243-7.
- MATSUMOTO, E. D., KONDRASKE, G. V., OGAN, K., JACOMIDES, L., WILHELM, D. M., PEARLE, M. S. & CADEDDU, J. A. 2006a. Assessment of basic human performance resources predicts performance of ureteroscopy. *Am J Surg*, 191, 817-20.
- MATSUMOTO, E. D., PACE, K. T. & RJ, D. A. H. 2006b. Virtual reality ureteroscopy simulator as a valid tool for assessing endourological skills. *Int J Urol*, 13, 896-901.
- MICHEL, M. S., KNOLL, T., KOHRMANN, K. U. & ALKEN, P. 2002. The URO Mentor: development and evaluation of a new computer-based interactive training system for virtual life-like simulation of diagnostic and therapeutic endourological procedures. *BJU Int*, 89, 174-7.
- MISHRA, S., KURIEN, A., GANPULE, A., MUTHU, V., SABNIS, R. & DESAI, M. 2010a. Percutaneous renal access training: content validation comparison between a live porcine and a virtual reality (VR) simulation model. *BJU Int*, 106, 1753-6.

- MISHRA, S., KURIEN, A., GANPULE, A., VEERAMANI, M., SABNIS, R. B. & DESAI, M. 2010b. Face and content validity of transurethral resection of prostate on Uro Trainer: is the simulation training useful? *J Endourol*, 24, 1839-43.
- MISHRA, S., KURIEN, A., PATEL, R., PATIL, P., GANPULE, A., MUTHU, V., SABNIS, R. B. & DESAI, M. 2010c. Validation of virtual reality simulation for percutaneous renal access training. *J Endourol*, 24, 635-40.
- MISHRA, S., SHARMA, R., KUMAR, A., GANATRA, P., SABNIS, R. B. & DESAI, M. R. 2011. Comparative performance of high-fidelity training models for flexible ureteroscopy: Are all models effective? *Indian J Urol*, 27, 451-6.
- MOLINAS, C. R., BINDA, M. M., MAILOVA, K. & KONINCKX, P. R. 2004. The rabbit nephrectomy model for training in laparoscopic surgery. *Hum Reprod*, 19, 185-90.
- NOURELDIN Y, E. M., FAHMY N, CARRIER S, ELHILALI M, ANDONIAN S 2014. Incorporation of the GreenLight-SIM Simulator at the Annual Quebec Urology Objective Structured Clinical Examinations. *Urology (Supplement)*, S139.
- OGAN, K., JACOMIDES, L., SHULMAN, M. J., ROEHRBORN, C. G., CADEDDU, J. A. & PEARLE, M. S. 2004. Virtual ureteroscopy predicts ureteroscopic proficiency of medical students on a cadaver. *J Urol*, 172, 667-71.
- PAPATSORIS, A. G., SHAIKH, T., PATEL, D., BOURDOUMIS, A., BACH, C., BUCHHOLZ, N., MASOOD, J. & JUNAID, I. 2012. Use of a virtual reality simulator to improve percutaneous renal access skills: a prospective study in urology trainees. *Urol Int*, 89, 185-90.
- PERSOON, M. C., SCHOUT, B. M., MUIJTJENS, A. M., HENDRIKX, A. J., WITJES, J. A. & SCHERPBIER, A. J. 2010. The effect of a low-fidelity model on cystoscopic skill training: a single-blinded randomized controlled trial. *Simul Healthc*, 5, 213-8.
- PERSOON, M. C., VAN PUTTEN, K., MUIJTJENS, A. M., WITJES, J. A., HENDRIKX, A. J. & SCHERPBIER, A. J. 2011. Effect of distraction on the performance of endourological tasks: a randomized controlled trial. *BJU Int*, 107, 1653-7.
- PONIATOWSKI, L. H., WOLF, J. S., JR., NAKADA, S. Y., REIHSEN, T. E., SAINFORT, F. & SWEET, R. M. 2014. Validity and acceptability of a high-fidelity physical simulation model for training of laparoscopic pyeloplasty. *J Endourol*, 28, 393-8.
- RASHID, H. H., KOWALEWSKI, T., OPPENHEIMER, P., OOMS, A., KRIEGER, J. N. & SWEET, R. M. 2007. The virtual reality transurethral prostatic resection trainer: evaluation of discriminate validity. *J Urol*, 177, 2283-6.
- REICH, O., NOLL, M., GRATZKE, C., BACHMANN, A., WAIDELICH, R., SEITZ, M., SCHLENKER, B., BAUMGARTNER, R., HOFSTETTER, A. & STIEF, C. G. 2006. High-level virtual reality simulator for endourologic procedures of lower urinary tract. *Urology*, 67, 1144-8.
- SABBAGH, R., CHATTERJEE, S., CHAWLA, A., HOOGENES, J., KAPOOR, A. & MATSUMOTO, E. D. 2012. Transfer of laparoscopic radical prostatectomy skills from bench model to animal model: a prospective, single-blind, randomized, controlled study. *J Urol*, 187, 1861-6.
- SCHOUT, B. M., ANANIAS, H. J., BEMELMANS, B. L., D'ANCONA, F. C., MUIJTJENS, A. M., DOLMANS, V. E., SCHERPBIER, A. J. & HENDRIKX, A. J. 2010a. Transfer of cysto-urethroscopy skills from a virtual-reality simulator to the operating room: a randomized controlled trial. *BJU Int*, 106, 226-31; discussion 231.
- SCHOUT, B. M., BEMELMANS, B. L., MARTENS, E. J., SCHERPBIER, A. J. & HENDRIKX, A. J. 2009. How useful and realistic is the uro trainer for training transurethral prostate and bladder tumor resection procedures? *J Urol*, 181, 1297-303; discussion 1303.
- SCHOUT, B. M., MUIJTJENS, A. M., HENDRIKX, A. J., ANANIAS, H. J., DOLMANS, V. E., SCHERPBIER, A. J. & BEMELMANS, B. L. 2010b. Acquisition of flexible cystoscopy skills on a virtual reality simulator by experts and novices. *BJU Int*, 105, 234-9.
- SHAH, J. & DARZI, A. 2002. Virtual reality flexible cystoscopy: a validation study. *BJU Int*, 90, 828-32.
- SHAH, J., MONTGOMERY, B., LANGLEY, S. & DARZI, A. 2002. Validation of a flexible cystoscopy course. *BJU Int*, 90, 833-5.
- SHAMIM KHAN, M., AHMED, K., GAVAZZI, A., GOHIL, R., THOMAS, L., POULSEN, J., AHMED, M., JAYE, P. & DASGUPTA, P. 2013. Development and implementation of centralized simulation training: evaluation of feasibility, acceptability and construct validity. *BJU Int*, 111, 518-23.

- SHERGILL, I. S., SHAIKH, T., ARYA, M. & JUNAID, I. 2008. A training model for suprapubic catheter insertion: the UroEmerge suprapubic catheter model. *Urology*, 72, 196-7.
- SINGAL, A., HALVERSON, A., ROONEY, D. M., DAVIS, L. M. & KIELB, S. J. 2015. A validated low-cost training model for suprapubic catheter insertion. *Urology*, 85, 23-6.
- SORIA, F., MORCILLO, E., SANZ, J. L., BUDIA, A., SERRANO, A. & SANCHEZ-MARGALLO, F. M. 2014. Description and validation of realistic and structured endourology training model. *Am J Clin Exp Urol*, 2, 258-65.
- SWEET, R., KOWALEWSKI, T., OPPENHEIMER, P., WEGHORST, S. & SATAVA, R. 2004. Face, content and construct validity of the University of Washington virtual reality transurethral prostate resection trainer. *J Urol*, 172, 1953-7.
- TEBER, D., GUVEN, S., YAYCIOGLU, O., UGURLU, O., SANLI, O., GOZEN, A. S. & RASSWEILER, J. 2010. Single-knot running suture anastomosis (one-knot pyeloplasty) for laparoscopic dismembered pyeloplasty: training model on a porcine bladder and clinical results. *Int Urol Nephrol*, 42, 609-14.
- TUNITSKY, E., MURPHY, A., BARBER, M. D., SIMMONS, M. & JELOVSEK, J. E. 2013. Development and validation of a ureteral anastomosis simulation model for surgical training. *Female Pelvic Med Reconstr Surg*, 19, 346-51.
- VENEZIANO, D., SMITH, A., REIHSEN, T., SPEICH, J. & SWEET, R. M. 2015. The SimPORTAL fluoro-less C-arm trainer: an innovative device for percutaneous kidney access. *J Endourol*, 29, 240-5.
- VOLPE, A., AHMED, K., DASGUPTA, P., FICARRA, V., NOVARA, G., VAN DER POEL, H. & MOTTRIE, A. 2014. Pilot Validation Study of the European Association of Urology Robotic Training Curriculum. *Eur Urol*.
- WATTERSON, J. D., BEIKO, D. T., KUAN, J. K. & DENSTEDT, J. D. 2002. Randomized prospective blinded study validating acquisition of ureteroscopy skills using computer based virtual reality endourological simulator. *J Urol*, 168, 1928-32.
- WHITE, M. A., DEHAAN, A. P., STEPHENS, D. D., MAES, A. A. & MAATMAN, T. J. 2010. Validation of a high fidelity adult ureteroscopy and renoscopy simulator. *J Urol*, 183, 673-7.
- WIJN, R. P., PERSON, M. C., SCHOUT, B. M., MARTENS, E. J., SCHERPBIER, A. J. & HENDRIKX, A. J. 2010. Virtual reality laparoscopic nephrectomy simulator is lacking in construct validity. *J Endourol*, 24, 117-22.
- WILHELM, D. M., OGAN, K., ROEHRBORN, C. G., CADEDDU, J. A. & PEARLE, M. S. 2002. Assessment of basic endoscopic performance using a virtual reality simulator. *J Am Coll Surg*, 195, 675-81.
- ZHANG, Y., LIU, J. S., WANG, G., YU, C. F., ZHU, H. & NA, Y. Q. 2013. Effectiveness of the UroMentor virtual reality simulator in the skill acquisition of flexible cystoscopy. *Chin Med J (Engl)*, 126, 2079-82.
- ZHANG, Y., OU, T. W., JIA, J. G., GAO, W., CUI, X., WU, J. T. & WANG, G. 2008. Novel biologic model for percutaneous renal surgery learning and training in the laboratory. *Urology*, 72, 513-6.
- ZHU, H., ZHANG, Y., LIU, J. S., WANG, G., YU, C. F. & NA, Y. Q. 2013. Virtual reality simulator for training urologists on transurethral prostatectomy. *Chin Med J (Engl)*, 126, 1220-3.

SUPPLEMENTARY TABLES

Table 1: Modified Levels of evidence classification for validation studies, adapted from Oxford Centre for Evidence-Based Medicine classification by the European Association of Endoscopic Surgeons (Carter et al., 2005).

Level of Evidence	Criteria
1a	Systematic reviews (meta-analysis) containing at least some trial of level 1b evidence, in which results of separate, independently controlled trials are consistent
1b	Randomised controlled trial of good quality and of adequate sample size (power calculations)
2a	Randomised trials of reasonable quality and/or of inadequate sample size
2b	Nonrandomised trials, comparative research (parallel cohort)
2c	Nonrandomised trials, comparative research (historical cohort, literature controls)
3	Nonrandomised, non-comparative trials, descriptive research
4	Expert opinions, including the opinion of Work Group members

Table 2: Levels of recommendation for training models, adapted from Oxford Centre for Evidence-Based Medicine classification by the European Association of Endoscopic Surgeons (Carter et al., 2005).

Level of Recommendation	Criteria
1	Based on one systematic review (1a) or at least two independently conducted research projects classified as 1b
2	Based on at least two independently conducted research projects classified as level 2a or 2b, within concordance
3	Based on one independently conducted research project level 2b, or at least two trials of level 3, within concordance
4	Based on one trial at level 3 or multiple expert opinions, including the opinions of Work Group members (e.g. level 4)

CARTER, F. J., SCHIJVEN, M. P., AGGARWAL, R., GRANTCHAROV, T., FRANCIS, N. K., HANNA, G. B., JAKIMOWICZ, J. J., PROGRAMMES, W. G. F. E., IMPLEMENTATION OF, S. & SKILLS, T. 2005. Consensus guidelines for validation of virtual reality surgical simulators. *Surgical endoscopy*, 19, 1523-1532.