King's Research Portal

DOI:

[10.1113/EP086867](https://doi.org/10.1113/EP086867)

Document Version

Peer reviewed version

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

Bruce, R. M., Phan, P. A., Pacpaco, E., Rahman, N. M., & Farmery, A. (2018). The inspired sinewave technique: a novel method to measure lung volume and ventilatory heterogeneity. *Experimental Physiology*, 738. <https://doi.org/10.1113/EP086867>

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Experimental Physiology

<https://ep.msubmit.net>

EP-RP-2017-086867R1

Title: THE INSPIRED SINEWAVE TECHNIQUE: A NOVEL METHOD TO MEASURE LUNG VOLUME AND VENTILATORY HETEROGENEITY

Authors: Richard M Bruce
Phi Anh Phan
Edmund Pacpaco
Najib Rahman
Andrew Farmery

Author Conflict: No competing interests declared

Running Title: The inspired sinewave technique: a novel test of lung function

Abstract: The Inspired Sinewave Technique (IST) is a novel method which can provide simple, non-invasive cardiopulmonary measurements. Over successive tidal-breaths the concentration of a tracer gas (i.e. nitrous oxide, N₂O) is sinusoidally modulated in inspired air. Using a single-compartment tidal-ventilation lung model, the resulting amplitude/phase of the expired sinewave allows estimation of end-expired lung volume (ELV), pulmonary blood flow and three indices for ventilatory heterogeneity (VH; ELV₁₈₀/FRC_{pleth}, ELV₁₈₀/FRC_{pred} and ELV₆₀/ELV₁₈₀). This investigation aimed to determine: the repeatability and agreement of ELV with FRC_{pleth}, and, as normal ageing results in well-established changes in pulmonary structure and function, whether the IST estimates of ELV and VH are age dependent. 48 healthy never-smoker participants (20-

86 years) underwent traditional pulmonary function testing: (e.g. spirometry, body plethysmography) and the IST test which consisted of 4 minutes of quiet breathing through a facemask while inspired N₂O concentrations are oscillated in a sinewave pattern with a fixed mean (4%) and amplitude (3%) and a period of either 180 seconds or 60 seconds. ELV^{180}/FRC^{pleth} and ELV^{180}/FRC_{pred} were age dependent (average decreases of 0.58% and 0.48% per year) suggesting an increase in VH with advancing age. ELV showed a mean bias of -1.09L vs. FRC_{pleth} , but when normalised for the effects of age this bias reduced to -0.35L. The IST test has potential to provide clinically useful information, (e.g. for mechanically ventilated patients), but these findings suggest that the increases in VH with healthy ageing must be accounted for in clinical investigations.

New Findings: We present a new non-invasive medical technology - the inspired sinewave technique - which involves inhaling sinusoidally fluctuating concentrations of a tracer gas. The technique requires only passive patient cooperation, and can monitor different cardiorespiratory variables such as end expired lung volume, ventilatory heterogeneity and pulmonary blood flow. In this article we demonstrate that the measurements of end expired lung volume are repeatable and accurate, in comparison to whole body plethysmography, and the technique is sensitive to the changes in ventilatory heterogeneity associated advancing ageing. As such, it has the potential to provide clinically valuable information.

Dual Publication: No

Funding: DH | National Institute for Health Research (NIHR): Richard M Bruce, Phi Anh Phan, Andrew Farmery, HMR00780

THE INSPIRED SINEWAVE TECHNIQUE: A NOVEL METHOD TO MEASURE LUNG VOLUME AND VENTILATORY HETEROGENEITY

Dr Richard M Bruce

School of Basic and Medical Sciences, Shepard's House, Guy's Campus, Kings College
London, London, UK SE1 1UL, Richard.bruce@kcl.ac.uk, +44 020 7848 6324

Dr Phi Anh Phan

Nuffield department of Clinical Neurosciences, University of Oxford, Oxford, UK

Mr Edmund Pacpaco

Oxford University Hospitals NHS Foundation Trust, Oxford Centre for Respiratory Medicine
Oxford, UK

Prof. Najib M Rahman

Oxford University Hospitals NHS Foundation Trust, Oxford Centre for Respiratory Medicine
Oxford, UK

Prof. Andrew Farmery

Nuffield department of Clinical Neurosciences, University of Oxford, Oxford, UK

Running title: The inspired sinewave technique: a novel test of lung function

Word count: 3222

References: 36

Key words: Respiratory physiology, Pulmonary function testing, Ageing

NEW FINDINGS

We present a new non-invasive medical technology - the inspired sinewave technique - which involves inhaling sinusoidally fluctuating concentrations of a tracer gas. The technique requires only passive patient cooperation, and can monitor different cardiorespiratory variables such as end expired lung volume, ventilatory heterogeneity and pulmonary blood flow.

In this article we demonstrate that the measurements of end expired lung volume are repeatable and accurate, in comparison to whole body plethysmography, and the technique is sensitive to the changes in ventilatory heterogeneity associated advancing ageing. As such, it has the potential to provide clinically valuable information.

ABSTRACT

The Inspired Sinewave Technique (IST) is a novel method which can provide simple, non-invasive cardiopulmonary measurements. Over successive tidal-breaths the concentration of a tracer gas (i.e. nitrous oxide, N₂O) is sinusoidally modulated in inspired air. Using a single-compartment tidal-ventilation lung model, the resulting amplitude/phase of the expired sinewave allows estimation of end-expired lung volume (ELV), pulmonary blood flow and three indices for ventilatory heterogeneity (VH; ELV_{180}/FRC_{pleth} , ELV_{180}/FRC_{pred} and ELV_{60}/ELV_{180}).

This investigation aimed to determine: the repeatability and agreement of ELV with FRC_{pleth} , and, as normal ageing results in well-established changes in pulmonary structure and function, whether the IST estimates of ELV and VH are age dependent.

48 healthy never-smoker participants (20-86 years) underwent traditional pulmonary function testing: (e.g. spirometry, body plethysmography) and the IST test which consisted of 4 minutes of quiet breathing through a facemask while inspired N₂O concentrations are oscillated in a sinewave pattern with a fixed mean (4%) and amplitude (3%) and a period of either 180 seconds or 60 seconds.

ELV_{180}/FRC_{pleth} and ELV_{180}/FRC_{pred} were age dependent (average decreases of 0.58% and 0.48% per year) suggesting an increase in VH with advancing age. ELV showed a mean bias of -1.09L vs. FRC_{pleth} , but when normalised for the effects of age this bias reduced to -0.35L. The IST test has potential to provide clinically useful information, (e.g. for mechanically ventilated patients), but these findings suggest that the increases in VH with healthy ageing must be accounted for in clinical investigations.

INTRODUCTION

Traditional pulmonary function tests, such as spirometry and whole body plethysmography, are a critical aspect of the diagnosis, characterisation, and management of lung disease (Pellegrino et al., 2005). However, they require specific and often unfamiliar respiratory manoeuvres, and so are unsuitable for critically ill or sedated patients, and their reproducibility is reduced in both elderly and paediatric populations (Bellia et al., 2000, Beydon et al., 2007). The inspired sinewave technique (IST) is a novel method of measuring, or continuously monitoring, cardiopulmonary function which requires only passive patient cooperation (Clifton et al., 2013, Farmery, 2008, Phan et al., 2015). Therefore, the technology can provide simple, non-invasive cardiopulmonary measurements in both the outpatient and critical care settings.

Over successive tidal breaths the concentration of a tracer gas (e.g. nitrous oxide, N_2O) is sinusoidally modulated in inspired air. The amplitude and phase of the expired sinewave are altered by the pulmonary ventilation and blood flow (if the tracer gas is soluble) and distorted further by ventilatory heterogeneity (VH). A mathematical model of the lung processes flow and concentration data and recovers values for cardio-respiratory variables such as end-expired lung volume (ELV), pulmonary blood flow ($\dot{Q}_p$) and, in addition, indices for VH can be calculated.

The technique originated from the work of Zwart et al. (Zwart et al., 1978, Zwart et al., 1976) and was extended by Hahn et al. (Hahn et al., 1993, Hahn, 1996) by using patient safe gases such as O_2 and low concentrations of N_2O , and developing a more realistic tidal ventilation lung model (Gavaghan and Hahn, 1996, Whiteley et al., 2001). Preliminary investigations in animals and human participants have since shown that the technique has the potential to

provide useful measures of cardiorespiratory function(Williams et al., 1998, Clifton et al., 2013, Farmery, 2008). In this investigation we firstly aimed to: 1) determine the accuracy of the IST measurement of ELV in healthy participants, by examining its agreement with FRC measured via whole body plethysmography (FRC_{pleth}), and 2) examine the repeatability of the ELV measurement. Moreover, as healthy ageing results in several alterations in pulmonary structure and function – such as increases in lung compliance, the enlargement of airspaces and increases in VH(Turner et al., 1968, Gillooly and Lamb, 1993, Verbanck et al., 2012) – this investigation also aimed to assess whether the IST estimates ELV and VH are age dependent.

MATERIALS AND METHODS

Ethical approval

All participants received verbal and written information regarding the protocol and experimental procedures before giving their written informed consent. The study's protocol was approved by an NHS ethical committee (16/SC/0057, ethics protocol 1.0) and conforms to the *Declaration of Helsinki, 2013*, except for registration in a database. 48 never-smoker participants volunteered for the study – characteristics and pulmonary function test data are shown in Table 1. The age of participating subjects ranged between 20-86, and all were defined as healthy through clinical screening with the following criteria: no childhood/past medical history of respiratory disease, no history of respiratory symptoms suggestive of disease, no upper respiratory tract infections in the previous 8 weeks, and no history of smoking.

N	48	
Male/Female	26/22	
	Mean (SD)	% Pred (SD)
Age (years)	53.5 (22)	-
Height (m)	1.7 (0.1)	-
Weight (kg)	71.5 (12)	-
BMI (kg/m²)	24.9 (4.2)	-
FEV₁ (L)	3.3 (1)	102.4 (17.1)
FVC (L)	4.3 (1.3)	105.6 (18.5)
FEV₁%FVC	77.8 (6.9)	96.5 (7.5)
TLC (L)	6.3 (1.3)	105.4 (13.2)
RV (L)	2.2 (0.8)	104.4 (20.7)
FRC (L)	3.4 (0.8)	107.5 (16.6)
T_{LCO}	7.8 (2.6)	98.7 (15.5)
K_{CO}	1.5 (0.3)	94.6 (16)
VA/TLC	0.87 (0.05)	-
ELV₁₈₀	2.3 (0.8)	-
ELV₆₀	1.5 (0.5)	-
Q_P	4.9 (1.7)	-

Table 1. Participant characteristics and pulmonary function test measurements. SD, standard deviation; BMI, body mass index; FEV₁, forced expiratory volume in one second; FVC, forced vital capacity; TLC, total lung capacity; residual volume; FRC, functional residual capacity; RV, residual volume; T_{LCO}, transfer factor for carbon monoxide; K_{CO}, transfer coefficient for carbon monoxide; VA/TLC, ratio between a single-breath helium dilution TLC measurement (VA) and TLC from whole body plethysmography; ELV₁₈₀ & ELV₆₀, IST estimate of end-expired lung volume measured using a

sinewave with a 180 second or 60 second period respectively; Q_p , IST estimate of pulmonary blood flow. % predicted values obtained via reference values: Quanjer et al 1993, 2012.

Study protocol

All procedures for each participant were performed on the same day within the Respiratory Medicine Department of the Churchill Hospital – Oxford University Hospitals NHS Foundation Trust. Prior to testing, a medical history and anthropometric data were recorded. Participants then underwent traditional pulmonary function testing: spirometry, body plethysmography and the single breath test of carbon monoxide uptake (Jaeger MasterScope Body, Carefusion); followed by the IST test.

IST test

Participants were seated and asked to breathe quietly through a face mask connected to a mainstream infrared N_2O and CO_2 sensor (SquareOne Technology) and an ultrasonic flowmeter (VenThor – 22/2A) – see figure 1 for a schematic representation of the set up. At the start of each inhalation a small quantity of N_2O was injected into the participants' inspired gas by a mass flow controller (Alicat Scientific, Inc., USA). The volume injected was proportional to each breath's inspiratory flow, and over successive breaths the concentration of inspired N_2O oscillates in a sinewave pattern around a set mean (4%) with a predetermined amplitude (3%) and frequency (60 sec or 180 sec period). The inspired N_2O fraction can therefore be defined by the following:

$$F_I(t) = \bar{F}_I + \Delta F_I \sin\left(\frac{2\pi}{T}t + \Phi\right) \quad \text{eq. 1}$$

where $\bar{F}_I$ and ΔF_I are the mean and amplitude of the N₂O sinusoid respectively, T is the period, and Φ is the phase.

In the absence of venous recirculation of the N₂O sinewave, which is known to be negligible at the periods used in the current study, i.e. <5min, (Hahn et al., 1993, Gavaghan and Hahn, 1995) forcing sinusoidal inspired N₂O concentrations results in the end-tidal (i.e. alveolar) N₂O concentrations to also oscillate sinusoidally. Figure 2 shows a typical data set recorded from the IST test.

Using a single-compartment tidal ventilation model of the lung, the resulting amplitude and phase of the expired sinewave allows the estimation of $\dot{Q}_P$ and alveolar volume (V_A). The sum of V_A and airways dead space (V_D), as measured via the Bohr technique (Phan et al., 2017), from the IST estimate of end-expired lung volume (ELV). The period (in seconds) of the inspired N₂O sinewave used to estimate ELV is signified by the subscript – i.e. ELV₁₈₀ or ELV₆₀. Modelling and preliminary empirical data has shown that ELV₁₈₀ has the closest agreement with FRC_{pleth}. The lung model and estimation steps have been described elsewhere (Clifton et al., 2013, Phan et al., 2015, Gavaghan and Hahn, 1996), a summary of which is given in the appendix.

Indices of ventilation heterogeneity (VH)

Three potential indices are proposed:

- 1) ELV₁₈₀/FRC_{pleth}
- 2) ELV₁₈₀/FRC_{pred}
- 3) ELV₆₀/ELV₁₈₀

where lower values suggest greater VH and higher values towards 1 suggest homogeneity.

The first two indices rely on the nature of *single compartment* models used to estimate lung volume (ELV). For the IST, ELV measurements in a uni-compartmental lung are a predictable function of ventilation, the sinewave period, and the attenuation of the expired sinewave as described by equation 1, and errors (underestimations) in the calculated volume suggests a degree of VH(Whiteley et al 2001). As such, increasing heterogeneity will result in greater disparity between ELV_{180} and FRC_{pleth} , or that estimated via prediction equations (FRC_{pred}):

$$Male FRC_{pred} = 2.34 \times H + 0.01 \times A - 1.09$$

$$Female FRC_{pred} = 2.24 \times H + 0.001 \times A - 1$$

where H is standing height in metres and A is age in years(14)

The third index relies on ELV measurements becoming more dependent on sinewave period with increasing VH. In a homogeneous (uni-compartmental) lung the ELV measurement should be unaffected by the period used. However, preliminary modelling and experimentation has uncovered that where VH exists, estimated cardiopulmonary variables become period dependent(Whiteley et al., 2001). Therefore, the ratio of two ELV measurements made using different sinewave periods (ELV_{60} and ELV_{180}) may estimate the degree of VH.

In addition, a further simple index (VA/TLC) of gas mixing inefficiency and VH has been calculated, which is the ratio between a single-breath helium dilution TLC measurement (VA) and a TLC measurement from whole body plethysmography (Cotes et al. 2006).

IST test protocol

Participants were seated and rested for 5 minutes prior to each IST test. For the duration of the IST test participants are asked to breathe quietly through a face mask, sealed with no

leaks, held up in front of them by an adjustable articulating arm. All participants performed two IST tests – 3 minutes of inhaling a constant concentration of N₂O (4%) followed by 4 minutes of forcing an inspired N₂O sinewave (mean = 4%, amplitude = 3%) at periods of 180 sec and 60 sec. Participants then repeated these two tests after a 15 minute interval.

Statistical analysis

All statistical analysis was conducted using Microsoft Excel (2016) and a standard statistical package (SPSS, Chicago, IL, USA). The agreement between FRC_{pleth} and ELV₁₈₀ was assessed using Bland-Altman analysis. The repeatability of duplicate ELV₁₈₀ measurements were assessed using linear regression analysis, Bland-Altman analysis and the coefficient of variation (standard deviation/mean x 100). The first and second ELV₁₈₀ measurements were also compared using a student's paired t-test with statistical significance taken as p<0.05. Multiple stepwise regression analysis was performed on the three VH indices: ELV/FRC_{pleth}, ELV/FRC_{pred}, ELV₆₀/ELV₁₈₀, and independent variables included age, sex, height, and FRC_{pleth}. Multiple stepwise regression analysis was also performed for ELV₁₈₀ and FRC_{pleth}, and independent variables included age, sex and height. The statistical significance level for retention was set at p<0.05. When sex was uncovered as a significant factor for the ELV₁₈₀ and FRC_{pleth} measurements, multiple regression analysis was repeated for each sex. Pearson's correlation coefficient was also calculated for the three IST indices of VH and pulmonary function test measurements: FEV₁, FEV₁/FVC and VA/TLC.

RESULTS

Indices of ventilation heterogeneity

Figures 3A and 3B show that ELV₁₈₀/FRC_{pleth} and ELV₁₈₀/FRC_{pred} decrease as a function of age, suggesting an increase in VH. The other assessed independent variables (sex, height and

FRC_{pleth}) did not reach statistical significance in the multiple regression analysis for all three heterogeneity indices, and although age was a significant contributor to the regressions of ELV₁₈₀/FRC_{pleth} and ELV₁₈₀/FRC_{pred} (p<0.05) it did not reach significance for ELV₆₀/ELV₁₈₀ (figure 3C). Regression equations for these VH indices are shown in table 2, and their correlation with pulmonary function test measurements are shown in table 3.

	Regression equations	Adjusted R ²
VH indices		
ELV ₁₈₀ /FRC _{pleth}	-0.0048 x A + 0.9427	0.62
ELV ₁₈₀ /FRC _{pred}	-0.0041 x A + 0.9478	0.28
ELV ₆₀ /ELV ₁₈₀	-	-
Lung volumes		
ELV ₁₈₀		
Men	4.956 X H - 0.012 x A - 5.325	0.71
Women	4.292 X H - 0.007 x A - 4.794	0.79
FRC _{pleth}		
Men	5.635 X H + 0.008 x A - 4.989	0.71
Women	4.989 X H + 0.013 x A - 5.928	0.69

Table 2. Regression equations for IST indices of VH (ELV₁₈₀/FRC_{pleth}, ELV₁₈₀/FRC_{pred}, ELV₆₀/ELV₁₈₀), and for ELV₁₈₀ and FRC_{pleth}. H, Height (m); A, age(years).

	FEV ₁	FEV ₁ /FVC	VA/TLC
VH Indices			
ELV ₁₈₀ /FRC _{pleth}	0.63*	0.34*	0.56*
ELV ₁₈₀ /FRC _{pred}	0.64*	0.15	0.57*
ELV ₆₀ /ELV ₁₈₀	0.08	-0.02	0.04

Table 3. Pearson's correlation coefficients (r) for IST indices of VH (ELV₁₈₀/FRC_{pleth}, ELV₁₈₀/FRC_{pred}, ELV₆₀/ELV₁₈₀), with FEV₁, FEV₁/FVC and VA/TLC. *, p<0.05).

Lung volume (ELV₁₈₀): repeatability and agreement with FRC_{pleth}

The mean measured ELV₁₈₀ was 2.31 litres (± 0.75 ; range: 1.01-3.55). Bland-Altman analysis (figure 4) shows that the mean differences between repeated measurements was -0.05 litres (± 0.2), 1.29% of the mean, and the 95% limits of agreement were between -0.42 and 0.31 litres. The coefficient of variation for repeated ELV₁₈₀ measurements was 4.8%. Linear regression analysis between these measurements produced an R^2 of 0.94, $y = 0.983 + 0.0871x$ (figure 5). The first and second measurement were not statistically different ($p = 0.76$).

The mean measured FRC_{pleth} was 3.4 litres (0.85; range 2.11-5.19), with an average percent predicted value of 107.5% (16.6; range 67.4-120). Bland-Altman analysis (figure 6) shows that the mean differences between FRC_{pleth} and ELV₁₈₀ measurements was -1.08 litres (0.53), and the 95% limits of agreement were between -2.1 and -0.37 litres. Regression equations for ELV₁₈₀ and FRC_{pleth} from the current cohort of participants are shown in table 2. An age-normalised Bland-Altman plot is shown in figure 7.

DISCUSSION

Lung volume

If we accept FRC_{pleth} as the gold standard measure of end-expiratory lung volume, the current study suggests that the IST test underestimates lung volume by approximately 1.09L. Underestimations are expected as body plethysmography measures the volume of compressible gas in the thorax and inevitably includes volumes that are either located in regions that are poorly ventilated or that do not communicate with airways – including that within the abdominal cavity. This volume cannot be quantified by the IST test, or traditional dilution tests (Schaanning and Gulsvik, 1973), and so contributes to an inevitable bias.

However, the underestimation of ELV_{180} relative to FRC_{pleth} clearly has a component of age dependency. Multivariate linear regression analysis on the current cohort uncovered that FRC_{pleth} increases by approximately 0.01L per year of age and, interesting, ELV decreases by approximately -0.01L per year (Table 2). This implies that the accuracy of the IST estimate of lung volume reduces with age (and likely reflects increasing VH – see Ageing and Ventilatory Heterogeneity section). However, the agreement of ELV_{180} with FRC_{pleth} can be improved. Using the regression equations from the multivariate analysis, ELV_{180} and FRC_{pleth} can be normalised to that predicted of a 20-year-old for a given height and sex. Bland-Altman analysis of age-normalised ELV_{180} and FRC_{pleth} (figure 7) shows a much smaller mean bias of -0.35L (with 95% limits of agreement at $\pm 0.73L$). This underestimation of ELV_{180} is similar to that of helium dilution measurement of FRC_{pleth} (e.g. -0.3L; Schaanning and Gulsvik, 1973). As such, simply normalising the ELV_{180} measurement for age provides a more accurate measurement of end-expiratory lung volume.

The ELV measurement was very repeatable between tests separated by a 15 minute period, with a coefficient of variation of 4.77% - similar that of whole body plethysmography and helium dilution measurements of end-expired lung volume, which range from approximately 4-7% and 5-10% respectively (Hankinson et al., 1998). There was no effect of test order on the duplicated measurements as shown by Bland-Altman analysis of the first and second ELV measurements (mean bias of 0.05L, 95% limits of agreement of 0.36), and the failure of their means to reach statistically significant difference. The mean absolute percentage difference between the duplicated measurements was 7.3% (± 7.6). This complies with the current standard guidelines for acceptable differences between duplicated helium dilution and nitrogen washout measurements of lung volumes (Wanger et al., 2005) (10%).

Accurate and precise measurements of lung volume offer valuable information to respiratory clinicians, as alterations in lung mechanics occur in several pathologies (Pride and Macklem,

2011). In addition, the IST test might provide clinically useful information in the critical care setting; Continuous monitoring of ELV could be of great value in ventilated patients where positive end expiratory pressures have been applied, by reducing the risk of ventilator induced lung injury via volutrauma or atelectrauma (Slutsky and Ranieri 2013). #

However, whilst volume estimations are based upon a single compartment lung model, significant respiratory disease and VH will likely impair the accuracy of the technique. Indeed, multi-compartmental lung models could be used to provide better estimates of lung volume where VH exist (Whiteley et al. 2001), and further work in patients is required to examine this. However, an advantage of using uni-compartment models is that they can be more easily ‘inverted’ – where by recorded physiological data (e.g. the attenuation of the expired sinewave) can be inserted into them and recover estimations of cardiorespiratory parameters (e.g. ELV or $\dot{Q}_P$; Hahn & Farmery, 2003). Indeed, finding inverse solutions for these variables in the heterogeneous lung can be problematic when using more complex and mathematically flexible multi-compartment models (Whiteley et al. 2001). In addition, the physiological interpretation of such models can be challenging for physiologists and clinicians. Altogether, the practicality and value of using single compartment models have ensured their continued use (Farmery, 2008; Hahn & Farmery 2003).

Ageing and Ventilatory Heterogeneity

It is well established that ageing results in several changes in pulmonary structure and function, including an enlargement of airspaces and increases in airflow obstruction (Janssens et al., 1999), both of which likely contribute to greater degrees of VH. Single-breath and multi-breath nitrogen washout (SBNW and MBNW) tests assess the efficiency of ventilation distribution, where greater VH is reflected in steeper phase III slopes and delayed N₂

clearance respectively (Robinson et al., 2013). With advancing age, phase III of N₂ expirograms become steeper (Sandqvist and Kjellmer, 1960, Roberts et al., 1991), and the clearance of N₂ during a MBNW test becomes delayed – as demonstrated by the positive linear associations between age and LCI, S_{cond} and S_{acin} (Verbanck et al., 2012).

The degree of disparity between the IST measure of lung volume (ELV₁₈₀) and the *true lung* volume, might similarly provide a useful index of VH. This is because ELV measurements are a perfectly predictable function of ventilation, the sinewave period, and its attenuation, and errors in the calculated volume suggests a degree of VH (Whiteley et al 2001). This is comparable to a multi-breath nitrogen (N₂) washout test, where deviations in the N₂ washout curve from a perfectly exponential decay suggests a degree of VH. As such, the findings of the current study also suggest that healthy ageing is associated with increases in VH, as decreases in the ratio of ELV₁₈₀ to FRC_{pleth} and ELV₁₈₀ to FRC_{pred} with advancing age were observed. Indeed, there was a significant positive correlation between these IST indices of VH and VA/TLC, a simple measure of gas mixing inefficiency (Hansen, 2011), and also with measures of airflow obstruction (FEV₁ and FEV₁/FVC; table 3). The linear decrease in ELV₁₈₀/FRC_{pleth} occurred within the age range 20 to 85 years, and between these ages it declined by approximately 38% (an average of 0.58% per year). ELV₁₈₀/FRC_{pred} also decreased linearly and between the ages of 20 and 85, where it reduced by approximately 31% (an average of 0.48% per year). These changes in ELV₁₈₀/FRC_{pleth} and ELV₁₈₀/FRC_{pred} with advancing age are comparable to extrapolated increases in LCI, S_{cond} and S_{acin}, which increase by 0.3-0.7% per year between the ages 20-85 – assuming continued linearity (Verbanck et al., 2012).

The pulmonary changes associated with ageing have been collectively termed ‘senile-emphysema’ (Knudson, 1991, Janssens et al 1999). However, there are clear histological distinctions with *true* emphysema which also results in the destruction of alveolar walls and a

more heterogenous increases in airspace size suggesting significantly greater degrees of VH (Verbeken et al., 1992). Therefore, methods that quantify VH, such as the IST test, have potential as a clinically useful tool in the diagnosis and/or staging of chronic obstructive pulmonary disease; and further work in patients is required to examine this possibility. However, our findings from the current investigation suggest that the physiological increases in VH with healthy ageing must be accounted for in any clinical investigation.

We found no change in third IST index of VH (ELV_{60}/ELV_{180}) with age. Given the evidence that ageing can increase VH (Verbanck et al., 2012) and that our other indexes (ELV_{180}/FRC_{pleth} , ELV_{180}/FRC_{pred} and VA/TLC) demonstrate this occurring in the current group of participants, it is likely that this index has failed to detect existing differences in VH between participants. This could be a consequence of using an unsuitable pair of sinewave periods, or this index might not be sensitive enough to detect small changes in VH associated with age. This can be examined further by using the index to compare healthy participants with patients with obstructive lung diseases, a group known to have significantly increased VH (Aurora et al., 2004, Verbanck et al., 1998).

Summary

The IST test is a novel, non-invasive method of estimating cardiopulmonary function that requires only passive patient cooperation. The current study has demonstrated that the IST indices of VH, ELV_{180}/FRC_{pleth} and ELV_{180}/FRC_{pred} , are sensitive to age. The IST test's estimation of end expired lung volume (ELV_{180}) was very repeatable, and has close agreement with FRC_{pleth} when normalised for the effects of age. Altogether, the IST test has the potential to provide clinically useful information, and so further work with mechanically ventilated patients, and those with obstructive lung disease is warranted.

REFERENCES

aurora, P., Gustafsson, P., Bush, A., Lindblad, A., Oliver, C., Wallis, C. E. & Stocks, J.
(2004). Multiple breath inert gas washout as a measure of ventilation distribution in

children with cystic fibrosis. *Thorax*, 59, 1068.

<http://dx.doi.org/10.1136/thx.2004.022590>

Bellia, V., Pistelli, R., Catalano, F., Antonelli-Incalzi, R., Grassi, V., Melillo, G., Olivieri, D. & Rengo, F. (2000). Quality Control of Spirometry in the Elderly. *American Journal of Respiratory and Critical Care Medicine*, 161, 1094-1100.
<https://doi.org/10.1164/ajrccm.161.4.9810093>

Beydon, N., Davis, S. D., Lombardi, E., Allen, J. L., Arets, H. G. M., Aurora, P., Bisgaard, H., Davis, G. M., Ducharme, F. M., Eigen, H., Gappa, M., Gaultier, C., Gustafsson, P. M., Hall, G. L., Hantos, Z., Healy, M. J. R., Jones, M. H., Klug, B., Carlsen, K. C. L., Mckenzie, S. A., Marchal, F., Mayer, O. H., Merkus, P. J. F. M., Morris, M. G., Oostveen, E., Pillow, J. J., Seddon, P. C., Silverman, M., Sly, P. D., Stocks, J., Tepper, R. S., Vilozi, D. & Wilson, N. M. (2007). An Official American Thoracic Society/European Respiratory Society Statement: Pulmonary Function Testing in Preschool Children. *American Journal of Respiratory and Critical Care Medicine*, 175, 1304-1345. <https://doi.org/10.1164/rccm.200605-642ST>

Clifton, L., Clifton, D. A., Hahn, C. E. W. & Farmery, A. D. (2013). Assessment of lung function using a non-invasive oscillating gas-forcing technique. *Respiratory Physiology & Neurobiology*, 189, 174-182. <https://doi.org/10.1016/j.resp.2013.05.015>

Cotes, J.E., Chinn, D.J. & Miller, M.R. (2006) Lung Function. Physiology, measurement and application in medicine. Sixth Edition, Blackwell Publishing.

Farmery, A. (2008). *Interrogation of the Cardiopulmonary System with Inspired GasTension Sinusoids*.

Gavaghan, D. J. & Hahn, C. E. (1995). A mathematical evaluation of the alveolar amplitude response technique. *Respir Physiol*, 102, 105-20.
<https://doi.org/10.1016/j.resp.2013.05.015>

- Gavaghan, D. J. & Hahn, C. E. (1996). A tidal breathing model of the forced inspired inert gas sinewave technique. *Respir Physiol*, 106, 209-21. [https://doi.org/10.1016/S0034-5687\(96\)00066-7](https://doi.org/10.1016/S0034-5687(96)00066-7)
- Gillooly, M. & Lamb, D. (1993). Airspace size in lungs of lifelong non-smokers: effect of age and sex. *Thorax*, 48, 39-43. <http://dx.doi.org/10.1136/thx.48.1.39>
- Hahn, C. E. (1996). Oxygen respiratory gas analysis by sine-wave measurement: a theoretical model. *J Appl Physiol* (1985), 81, 985-97. <https://doi.org/10.1152/jappl.1996.81.2.985>
- Hahn, C. E., Black, A. M., Barton, S. A. & Scott, I. (1993). Gas exchange in a three-compartment lung model analyzed by forcing sinusoids of N₂O. *J Appl Physiol* (1985), 75, 1863-76. <https://doi.org/10.1152/jappl.1993.75.4.1863>
- Hahn, C. E. & Farmery, A. D. (2003) Gas exchange modelling: no more gills, please. *British journal of anaesthesia* 91: 2-15. <https://doi.org/10.1093/bja/aeg142>
- Hankinson, J. L., Stocks, J. & Peslin, R. (1998). Reproducibility of lung volume measurements. *Eur Respir J*, 11, 787-90. doi:10.1183/09031936.98.11030787
- Hansen, J. E. (2011) Pulmonary function testing and interpretation. Jaypee Brothers Medical Publishers. doi: 10.5005/jp/books/11277
- Janssens, J. P., Pache, J. C. & Nicod, L. P. (1999). Physiological changes in respiratory function associated with ageing. *Eur Respir J*, 13, 197-205. doi: 10.1034/j.1399-3003.1999.13a36.x
- Knudson, R. (1991). Physiology of the aging lung. In: CRYSTAL, J. & WEST, J. (eds.) *The lung*. New York: Raven Press.
- Pellegrino, R., Viegi, G., Brusasco, V., Crapo, R. O., Burgos, F., Casaburi, R., Coates, A., Van Der Grinten, C. P. M., Gustafsson, P., Hankinson, J., Jensen, R., Johnson, D. C., Macintyre, N., McKay, R., Miller, M. R., Navajas, D., Pedersen, O. F. & Wanger, J.

- (2005). Interpretative strategies for lung function tests. *European Respiratory Journal*, 26, 948-968. doi: 10.1183/09031936.05.00035205
- Phan, P. A., Farmery, A. & Hahn, C. (2015). *Method and Apparatus for Measurement of Cardiopulmonary Function*. European Patent Number EP 3122249 A1. US Patent number US20170100043 A1.
- Phan, P. A., Hahn, C. E. W. & Farmery, A. D. (2017). A modification of the Bohr method to determine airways deadspace for non-uniform inspired gas tensions. *Physiol Meas*, 38, N107-N117. <http://dx.doi.org/10.1088/1361-6579/aa70c0>
- Pride, N. B. & Macklem, P. T. (2011). Lung Mechanics in Disease. *Comprehensive Physiology*. John Wiley & Sons, Inc. doi: 10.1002/cphy.cp030337
- Quanjer PH, Stanojevic S, Cole TJ, Baur X, Hall GL, Culver B, Enright PL, Hankinson JL, Ip MS, Zheng J, and Stocks J. (2012) Multi-ethnic reference values for spirometry for the 3-95 year age range: the global lung function 2012 equations. *European Respiratory Journal* 2012. doi: 10.1183/09031936.00080312
- Quanjer, P. H., Tammeling, G. J., Cotes, J. E., Pedersen, O. F., Peslin, R. & Yernault, J.-C. (1993). Lung volumes and forced ventilatory flows - Report Working Party, Standardization of Lung Function Tests, European Community for Steel and Coal and European Respiratory Society. *European Respiratory Journal*, 6, 5-40. doi: 10.1183/09041950.005s1693
- Roberts, C. M., Macrae, K. D., Winning, A. J., Adams, L. & Seed, W. A. (1991). Reference values and prediction equations for normal lung function in a non-smoking white urban population. *Thorax*, 46, 643-50. <http://dx.doi.org/10.1136/thx.46.9.643>
- Robinson, P. D., Latzin, P., Verbanck, S., Hall, G. L., Horsley, A., Gappa, M., Thamrin, C., Arets, H. G., Aurora, P., Fuchs, S. I., King, G. G., Lum, S., Macleod, K., Paiva, M., Pillow, J. J., Ranganathan, S., Ratjen, F., Singer, F., Sonnappa, S., Stocks, J.,

- Subbarao, P., Thompson, B. R. & Gustafsson, P. M. (2013). Consensus statement for inert gas washout measurement using multiple- and single- breath tests. *Eur Respir J*, 41, 507-22. doi: 10.1183/09031936.00069712
- Sandqvist, L. & Kjellmer, I. (1960). Normal values for the single breath nitrogen elimination test in different age groups. *Scand J Clin Lab Invest*, 12, 131-5.
- Schaanning, C. G. & Gulsvik, A. (1973). Accuracy and precision of helium dilution technique and body plethysmography in measuring lung volumes. *Scand J Clin Lab Invest*, 32, 271-7. <https://doi.org/10.3109/00365517309082471>
- Slutsky , A. S. & Ranieri , V. M. (2013). Ventilator-Induced Lung Injury. *New England Journal of Medicine*, 369, 2126-2136. DOI: 10.1056/NEJMra1208707
- Turner, J. M., Mead, J. & Wohl, M. E. (1968). Elasticity of human lungs in relation to age. *J Appl Physiol*, 25, 664-71. <https://doi.org/10.1152/jappl.1968.25.6.664>
- Verbanck, S., Schuermans, D., Muylem, A. V., Melot, C., Noppen, M., Vincken, W. & Paiva, M. (1998). Conductive and Acinar Lung-zone Contributions to Ventilation Inhomogeneity in COPD. *American Journal of Respiratory and Critical Care Medicine*, 157, 1573-1577. <https://doi.org/10.1164/ajrccm.157.5.9710042>
- Verbanck, S., Thompson, B. R., Schuermans, D., Kalsi, H., Biddiscombe, M., Stuart-Andrews, C., Hanon, S., Van Muylem, A., Paiva, M., Vincken, W. & Usmani, O. (2012). Ventilation heterogeneity in the acinar and conductive zones of the normal ageing lung. *Thorax*, 67, 789-95. <http://dx.doi.org/10.1136/thoraxjnl-2011-201484>
- Verbeken, E. K., Cauberghs, M., Mertens, I., Clement, J., Lauweryns, J. M. & Van De Woestijne, K. P. (1992). The senile lung. Comparison with normal and emphysematous lungs. 1. Structural aspects. *Chest*, 101, 793-9. <https://doi.org/10.1378/chest.101.3.793>

- Wanger, J., Clausen, J. L., Coates, A., Pedersen, O. F., Brusasco, V., Burgos, F., Casaburi, R., Crapo, R., Enright, P., Van Der Grinten, C. P., Gustafsson, P., Hankinson, J., Jensen, R., Johnson, D., Macintyre, N., McKay, R., Miller, M. R., Navajas, D., Pellegrino, R. & Viegi, G. (2005). Standardisation of the measurement of lung volumes. *Eur Respir J*, 26, 511-22. doi: 10.1183/09031936.05.00035005
- Whiteley, J. P., Gavaghan, D. J. & Hahn, C. E. (2001). A tidal breathing model of the inert gas sinewave technique for inhomogeneous lungs. *Respir Physiol*, 124, 65-83.
- Williams, E. M., Sainsbury, M. C., Sutton, L., Xiong, L., Black, A. M., Whiteley, J. P., Gavaghan, D. J. & Hahn, C. E. (1998). Pulmonary blood flow measured by inspiratory inert gas concentration forcing oscillations. *Respir Physiol*, 113, 47-56. [https://doi.org/10.1016/S0034-5687\(00\)00185-7](https://doi.org/10.1016/S0034-5687(00)00185-7)
- Zwart, A., Bogaard, J. M., Jansen, J. R. & Versprille, A. (1978). A non-invasive determination of lung perfusion compared with the direct Fick method. *Pflugers Arch*, 375, 213-7. <https://doi.org/10.1007/BF00584246>
- Zwart, A., Seagrave, R. C. & Van Dieren, A. (1976). Ventilation-perfusion ratio obtained by a noninvasive frequency response technique. *J Appl Physiol*, 41, 419-24. <https://doi.org/10.1152/jappl.1976.41.3.419>

APPENDIX: SUMMARY OF THE INSPIRED SINEWAVE TECHNIQUE

Figure A1 Illustrates the single compartment model of the cardiopulmonary system. The lung consists of one dead space compartment V_D , and one lung compartment $V_A(t)$ that expands and contracts with tidal volume. The body has one compartment V_{body} , with $\dot{Q}_P$ flowing through the lung for gas exchange.

When a tracer gas (e.g. N_2O) is inhaled, it passes through the dead space into the lung. It then diffuses into arterial blood, and distributes to different parts of the body. After some transit time, the tracer gas travels back to the lung in mixed venous blood. Some will cross the blood-gas barrier into the lung, before eventually being exhaled in expired gas.

During an IST test, over successive breaths the inspired concentration of the tracer gas $F_I(t)$ follows a sinewave pattern. Consequently, after a short transient time the alveolar ($F_A(t)$), arterial ($F_a(t)$) and mixed venous ($F_{\bar{v}}(t)$) concentrations will all follow sinewave patterns.

The body compartment acts as a low-pass filter to diminish the amplitude of the mixed venous sinewave $F_{\bar{v}}(t)$. When the sinewave period is short enough (≤ 5 mins), the amplitude of $F_{\bar{v}}(t)$ is small enough to assume it is constant, and therefore $F_{\bar{v}}(t) = F_I^0$ in which F_I^0 is the mean of inspired sinewave $F_I(t)$.

The mass balance of the lung compartment for breath n^{th} therefore can be written as:

$$F_{E,n-1} \times V_A + F_{I,n} \times (V_{T,n} - V_D) + F_{E,n-1} \times V_D - \lambda \times \dot{Q}_P \times (F_{E,n} - F_{\bar{v}}) \times \Delta t_n = V_A \times F_{E,n} + V_{T,n} \times F_{E,n}$$

$$\Leftrightarrow V_A \times (F_{E,n} - F_{E,n-1}) + \lambda \times \dot{Q}_P \times (F_{E,n} - F_I^0) \times \Delta t_n = V_D \times (F_{E,n-1} - F_{I,n}) + V_{T,n} \times (F_{I,n} - F_{E,n})$$

in which:

V_A : the alveolar lung volume;

$\dot{Q}_P$: the pulmonary blood flow;

V_D : the deadspace, estimated from the Bohr method;

$F_{\bar{I},n}$: the mean inspired concentration of breath nth;

$F_{E,n-1}, F_{E,n}$: the end expired concentrations of breath (n-1) and n;

λ : the solubility of N₂O, 0.47;

$F_{\bar{v}}$: the mixed venous concentration, assumed to be equal to the mean of the inspired sinewave concentration at steady state F_I^0 ;

Δt_n : the duration of breath nth;

$V_{T,n}$: the tidal volume of breath nth;

First, V_D is estimated using the Bohr method applied to an N₂O signal:

$$V_D = V_T \frac{F_E - F_{\bar{E}}}{F_E - F_{\bar{I}}}$$

in which $F_{\bar{E}}$ is the mean expired concentration. To compensate for the error associated with the non-uniform inspired concentration, a modified Bohr method has been proposed improving the accuracy of the airway deadspace estimation(Phan et al., 2017). For a series of breaths, a set of linear equations can be established from eq (2) and solved for V_A and $\dot{Q}_P$. ELV can then be calculated as sum of the V_D and V_A . Theoretically, three consecutive breaths would be sufficient to construct the linear equations to find solutions for V_A and $\dot{Q}_P$. In practice, the use of all breaths within a complete sinewave period is recommended to improve repeatability. Interested readers can refer elsewhere(Phan et al., 2015) for further detail.

FIGURES

Figure 1. Schematic representation of the Inspired Sinewave Device. The participant inhales and exhales air through flow meter (FM) and gas sensor (GS). At the start of each inspiration a small quantity of N_2O is delivered into the participants inspired gas via a mass flow controller (MFC). A mathematical model of the lung processes flow and concentration data and recovers values for cardio-respiratory variables such as dead space, alveolar volume, pulmonary blood flow, and, in addition, an index for ventilation heterogeneity can be calculated.

Figure 2. A typical data set collected from one participant during a 4 minute IST test. The green line is the expired N_2O concentration measured by the mainstream infrared gas sensor. The blue and red crosses are the N_2O concentrations in inspired gas and end-tidal gas respectively. The blue and red lines are the inspired and expired N_2O sinewaves respectively.

Figure 3A-C. Scatterplot of ELV_{180}/FRC_{pleth} , ELV_{180}/FRC_{pred} , and ELV_{60}/ELV_{180} versus age with a linear regression line.

Figure 4. Bland-Altman plot showing the agreement of two repeated IST measurements of ELV ($ELV_{180(1)}$ vs $ELV_{180(2)}$). LOA, limits of agreement.

Figure 5. Linear regression analysis of the two repeated IST measurements of ELV_{180} ($ELV_{180(1)}$ vs. $ELV_{180(2)}$).

Figure 6. Bland-Altman plot showing the agreement between ELV_{180} and FRC_{pleth} . LOA, limits of agreement.

Figure 7. Bland-Altman plot showing the agreement between age-normalised ELV_{180} and FRC_{pleth} . Using the regression equations from multivariate analysis, ELV_{180} and FRC_{pleth} are

normalised to that predicted of a 20 year old for a given height and sex. LOA, limits of agreement.

Figure A.1: The single compartment model of the lung and circulatory system

AUTHORS CONTRIBUTIONS

Experiments were performed within the Respiratory Medicine Department of the Churchill Hospital – Oxford University Hospitals NHS Foundation Trust. RMB, PP, NR, AF contributed to the conception and design of the work; RMB, PP, EP, AF contributed to the acquisition, analysis and interpretation of the data; RMB, PP, EP, NR, AF contributed to the drafting of the work and/or revising it critically. All authors approved the final version of the manuscript and agree to be accountable for all aspects of the work in ensuring that questions related to the accuracy or integrity of any part of the work are appropriately investigated and resolved. All persons designated as authors qualify for authorship, and all those who qualify for authorship are listed.

FUNDING

The work was funded by a National Institute for Health Research grant (HMR00780).

ACKNOWLEDGEMENTS

The authors appreciate the time given up by the research participants, and the important technical assistance from the Clinical Scientists Ben Guan, Blake Marsh and Michaela Paish.

Gas Delivery system

LabVIEW
Controller

MFC

FM

GS

Participant

LabVIEW data
acquisition

MATLAB
analysis

Dead space

Lung Volume

Pulmonary blood flow

Data capture/analysis