



King's Research Portal

DOI:

[10.5281/zenodo.572579](https://doi.org/10.5281/zenodo.572579)

Document Version

Publisher's PDF, also known as Version of record

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

Bonnici, T. A., Charlton, P., Alastruey, J., Tarassenko, L., Watkinson, P., & Beale, R. (2014). Continuous Physiological Monitoring of Ambulatory Patients. In *MEC Annual Meeting and Bioengineering14 Programme and Abstracts* (pp. 38-38). MECbioeng14, Imperial College London. <https://doi.org/10.5281/zenodo.572579>

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.



Continuous Physiological Monitoring of Ambulatory Patients

T. Bonnici^{1,2} and P. Charlton^{1,2}, J. Alastruey¹, L. Tarassenko³, P.J. Watkinson⁴, R. Beale^{1,2}

¹King's College London ²Guy's and St Thomas' NHS Foundation Trust

³University of Oxford ⁴Oxford Biomedical Research Centre

Guy's and St Thomas' **NHS**
NHS Foundation Trust



Using wireless sensors for earlier detection of physiological derangements than routine observations.

For what proportion of patient stay can continuous ECG and pulse oximetry data be captured using wireless sensors?

222 patients were monitored wirelessly during their recovery from cardiac surgery on a general ward. ECG data were acquired for the majority of patient stay, whereas pulse oximetry data were acquired for less than a fifth of the time. Furthermore, over a third of patients requested to stop wearing sensors. Therefore, there is room for improvement to the design of wireless sensors, particularly the pulse oximetry component.

Method

A post-surgical cardiac ward was equipped with wireless sensors which transmitted data to a central monitor in real time. 222 patients consented to wear a sensor whilst recovering from cardiac surgery.

Conclusions

ECG data were acquired for the majority of patient stay, whereas pulse oximetry data were acquired for less than a fifth of the time. Furthermore, over a third of patients requested to stop wearing sensors. Therefore, there is room for improvement to the design of wireless sensors, particularly the pulse oximetry component.

Results

196 patients wore a sensor (22 did not stay on the ward, 3 requested not to wear a sensor, and 1 was not given a sensor for clinical reasons). **122 wore a sensor until discharge**, 66 requested to stop wearing the sensor early, 6 had the sensor removed for clinical reasons, and 2 were transferred elsewhere. The median length of ward stay was 4.5 days. The 196 patients wore sensors for 898 out of 1344 days. **ECG and pulse oximetry data were acquired from each patient for 62% (52) and 18% (41) of their stay respectively.**

Future Work

This trial aims to determine the proportion of physiological derangements detected by wireless sensors over an hour before standard monitoring. Two examples are shown above.

The dataset will also be used to design risk prediction algorithms.

Ambulatory patients are monitored intermittently


Clinical deteriorations of hospital patients must be recognised early to maintain patient safety and minimise treatment costs. Routine practice on UK wards is to measure physiological parameters intermittently every 4-6 hours to aid recognition of deteriorations. Thus there may be a prolonged period of deterioration prior to recognition.

Patients' tolerance of wireless sensors is unknown

In a previous assessment of patients' tolerance of wireless sensors, none acquired data consistently for 24 hours [1]. Therefore, wireless sensors which are routinely used in the NHS were chosen for this study. Whilst more cumbersome than state-of-the-art sensors, they may be more reliable.

Deteriorations may be detected earlier by continuous monitoring

Wireless sensors, combined with risk prediction algorithms, may enable earlier detection of deteriorations by continuously monitoring physiology. However, such systems depend on reliable data acquisition. We assessed the proportion of patient stay for which continuous ECG and pulse oximetry data could be captured using wireless sensors.



References

[1] Bonnici T *et al.* Testing of wearable monitors in a real-world hospital environment: what lessons can be learnt? In *Ninth international conference on wearable and implantable body sensor networks*, 2012. Washington DC: IEEE, 2012; 79-84.

Acknowledgements

This research was supported by the EPSRC [Grant EP/F058845/1], the National Institute for Health Research (NIHR) comprehensive Biomedical Research Centre at Guy's & St Thomas' NHS Foundation Trust, and the NIHR Oxford Biomedical Research Centre Programme. The views expressed are those of the authors and not necessarily those of the EPSRC, the NHS, the NIHR or the Department of Health.

Contact

T. Bonnici and P. Charlton contributed equally to this work:
Timothy.Bonnici [at] ndm.ox.ac.uk
Peter.Charlton [at] gstt.nhs.uk